

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN120522\
 Data File : VN075559.D
 Acq On : 05 Dec 2022 15:19
 Operator : JC\MD
 Sample : N5809-01
 Misc : 5.00g/10mL/100uL/5.00mL/MSVOA_N/MEOH
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 HOWE-DRUM-A-1129

Quant Time: Dec 06 03:25:00 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N112122W.M
 Quant Title : SW846 8260
 QLast Update : Wed Nov 23 00:18:03 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.224	168	158358	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.100	114	254494	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.870	117	214529	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	94987	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.582	65	50825	59.791	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	119.580%
35) Dibromofluoromethane	8.171	113	63199	63.216	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	126.440%#
50) Toluene-d8	10.565	98	131395	58.354	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	116.700%#
62) 4-Bromofluorobenzene	12.853	95	90587	54.505	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	109.020%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN120522\
Data File : VN075559.D
Acq On : 05 Dec 2022 15:19
Operator : JC\MD
Sample : N5809-01
Misc : 5.00g/10mL/100uL/5.00mL/MSVOA_N/MEOH
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
HOWE-DRUM-A-1129

Quant Time: Dec 06 03:25:00 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N112122W.M
Quant Title : SW846 8260
QLast Update : Wed Nov 23 00:18:03 2022
Response via : Initial Calibration

