

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN120522\
 Data File : VN075565.D
 Acq On : 05 Dec 2022 17:43
 Operator : JC\MD
 Sample : N5793-02
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 B-40-SB01

Quant Time: Dec 06 03:26:04 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N112122W.M
 Quant Title : SW846 8260
 QLast Update : Wed Nov 23 00:18:03 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.230	168	172252	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	305831	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.871	117	279861	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	116650	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.582	65	53117	57.447	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	= 114.900%	
35) Dibromofluoromethane	8.177	113	66151	55.061	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 110.120%	
50) Toluene-d8	10.571	98	140654	51.980	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	= 103.960%	
62) 4-Bromofluorobenzene	12.853	95	103297	51.720	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	= 103.440%	
Target Compounds						
					Qvalue	
16) Acetone	4.442	43	30054	34.847	ug/l #	88
20) Methylene Chloride	5.300	84	7186	2.904	ug/l #	67
37) Ethyl Acetate	7.565	43	20531	8.703	ug/l	99
43) Isopropyl Acetate	8.694	43	96639	25.521	ug/l #	91

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN120522\
Data File : VN075565.D
Acq On : 05 Dec 2022 17:43
Operator : JC\MD
Sample : N5793-02
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 15 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
B-40-SB01

Quant Time: Dec 06 03:26:04 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N112122W.M
Quant Title : SW846 8260
QLast Update : Wed Nov 23 00:18:03 2022
Response via : Initial Calibration

