

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN120522\
 Data File : VN075567.D
 Acq On : 05 Dec 2022 18:31
 Operator : JC\MD
 Sample : N5844-02
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 MH-17

Quant Time: Dec 06 03:26:22 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N112122W.M
 Quant Title : SW846 8260
 QLast Update : Wed Nov 23 00:18:03 2022
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.230	168	168664	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	301866	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.871	117	278200	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	124859	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.577	65	49821	55.029	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	110.060%
35) Dibromofluoromethane	8.171	113	62835	52.988	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	105.980%
50) Toluene-d8	10.571	98	133621	50.030	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	100.060%
62) 4-Bromofluorobenzene	12.853	95	100150	50.803	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	101.600%
Target Compounds						
					Qvalue	
16) Acetone	4.442	43	14821	17.550	ug/l	90
20) Methylene Chloride	5.283	84	11496	5.577	ug/l #	80
37) Ethyl Acetate	7.565	43	22440	9.637	ug/l	98
43) Isopropyl Acetate	8.694	43	150508	40.269	ug/l #	81
52) Toluene	10.630	92	9244	1.812	ug/l	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN120522\
Data File : VN075567.D
Acq On : 05 Dec 2022 18:31
Operator : JC\MD
Sample : N5844-02
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 17 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
MH-17

Quant Time: Dec 06 03:26:22 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N112122W.M
Quant Title : SW846 8260
QLast Update : Wed Nov 23 00:18:03 2022
Response via : Initial Calibration

