

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN120522\
 Data File : VN075570.D
 Acq On : 05 Dec 2022 19:42
 Operator : JC\MD
 Sample : PB149407ZHE#03
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB149407ZHE#03

Quant Time: Dec 06 03:26:51 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N112122W.M
 Quant Title : SW846 8260
 QLast Update : Wed Nov 23 00:18:03 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.230	168	158560	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	279178	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.871	117	251506	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	100805	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.583	65	40488	47.570	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	95.140%
35) Dibromofluoromethane	8.177	113	49838	45.443	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	90.880%
50) Toluene-d8	10.571	98	104055	42.126	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	84.260%#
62) 4-Bromofluorobenzene	12.847	95	70566	38.705	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	77.400%#
Target Compounds						
					Qvalue	
16) Acetone	4.453	43	11354	14.302	ug/l	93
18) Methyl Acetate	5.048	43	4830	1.148	ug/l	99
20) Methylene Chloride	5.283	84	4914	1.820	ug/l	89
37) Ethyl Acetate	7.565	43	20598	9.565	ug/l	99
43) Isopropyl Acetate	8.694	43	131042	37.910	ug/l #	85

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN120522\
Data File : VN075570.D
Acq On : 05 Dec 2022 19:42
Operator : JC\MD
Sample : PB149407ZHE#03
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 20 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB149407ZHE#03

Quant Time: Dec 06 03:26:51 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N112122W.M
Quant Title : SW846 8260
QLast Update : Wed Nov 23 00:18:03 2022
Response via : Initial Calibration

