

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN120522\
 Data File : VN075572.D
 Acq On : 05 Dec 2022 20:30
 Operator : JC\MD
 Sample : PB149407ZHE#05
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 22 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB149407ZHE#05

Quant Time: Dec 06 03:27:10 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N112122W.M
 Quant Title : SW846 8260
 QLast Update : Wed Nov 23 00:18:03 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.230	168	155032	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	264705	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.871	117	243708	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	101327	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.582	65	42728	51.344	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	102.680%
35) Dibromofluoromethane	8.171	113	52913	50.885	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	101.780%
50) Toluene-d8	10.571	98	116898	49.913	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	99.820%
62) 4-Bromofluorobenzene	12.853	95	81007	46.861	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	93.720%
Target Compounds						
					Qvalue	
16) Acetone	4.442	43	11529	14.853	ug/l	99
18) Methyl Acetate	5.042	43	4510	1.036	ug/l	98
20) Methylene Chloride	5.294	84	4872	1.864	ug/l	89
37) Ethyl Acetate	7.571	43	21817	10.685	ug/l	99
43) Isopropyl Acetate	8.694	43	280425	85.561	ug/l #	70

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN120522\
Data File : VN075572.D
Acq On : 05 Dec 2022 20:30
Operator : JC\MD
Sample : PB149407ZHE#05
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 22 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB149407ZHE#05

Quant Time: Dec 06 03:27:10 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N112122W.M
Quant Title : SW846 8260
QLast Update : Wed Nov 23 00:18:03 2022
Response via : Initial Calibration

