

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN120522\
 Data File : VN075574.D
 Acq On : 05 Dec 2022 21:18
 Operator : JC\MD
 Sample : N5849-01
 Misc : 5.02g/10mL/100uL/5.00mL/MSVOA_N/MEOH
 ALS Vial : 24 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 TR-5-120222

Quant Time: Dec 06 03:27:27 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N112122W.M
 Quant Title : SW846 8260
 QLast Update : Wed Nov 23 00:18:03 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.224	168	170318	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.100	114	279169	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.870	117	231282	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	101670	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.582	65	45800	50.096	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	= 100.200%	
35) Dibromofluoromethane	8.171	113	58783	53.601	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 107.200%	
50) Toluene-d8	10.571	98	121671	49.259	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	= 98.520%	
62) 4-Bromofluorobenzene	12.853	95	82830	45.433	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	= 90.860%	
Target Compounds						Qvalue
18) Methyl Acetate	5.053	43	5799	1.440	ug/l	# 73

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN120522\
Data File : VN075574.D
Acq On : 05 Dec 2022 21:18
Operator : JC\MD
Sample : N5849-01
Misc : 5.02g/10mL/100uL/5.00mL/MSVOA_N/MEOH
ALS Vial : 24 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
TR-5-120222

Quant Time: Dec 06 03:27:27 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N112122W.M
Quant Title : SW846 8260
QLast Update : Wed Nov 23 00:18:03 2022
Response via : Initial Calibration

