

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN120522\
 Data File : VN075558.D
 Acq On : 05 Dec 2022 14:55
 Operator : JC\MD
 Sample : VN1205WBSD01
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VN1205WBSD01

Manual Integrations
 APPROVED

Reviewed By :Krupa Patel 12/06/2022
 Supervised By :Mahesh Dadoda 12/06/2022

Quant Time: Dec 06 03:24:44 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N112122W.M
 Quant Title : SW846 8260
 QLast Update : Wed Nov 23 00:18:03 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.230	168	167251	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	271743	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.870	117	246481	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	126617	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.582	65	47057	52.415	ug/l	0.00
Spiked Amount 50.000	Range 74 - 125		Recovery = 104.820%			
35) Dibromofluoromethane	8.177	113	60728	56.888	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery = 113.780%			
50) Toluene-d8	10.571	98	126181	52.481	ug/l	0.00
Spiked Amount 50.000	Range 86 - 113		Recovery = 104.960%			
62) 4-Bromofluorobenzene	12.853	95	97253	54.802	ug/l	0.00
Spiked Amount 50.000	Range 83 - 123		Recovery = 109.600%			
Target Compounds						
					Qvalue	
2) Dichlorodifluoromethane	2.136	85	23241	18.173	ug/l	98
3) Chloromethane	2.371	50	31002	19.482	ug/l	99
4) Vinyl Chloride	2.524	62	38482	19.188	ug/l	93
5) Bromomethane	2.948	94	33907	19.226	ug/l	96
6) Chloroethane	3.124	64	32085	21.113	ug/l	95
7) Trichlorofluoromethane	3.506	101	53705	18.744	ug/l	96
8) Diethyl Ether	3.971	74	21334	21.245	ug/l	99
9) 1,1,2-Trichlorotrifluo...	4.383	101	31450	19.884	ug/l	94
10) Methyl Iodide	4.600	142	48889	19.075	ug/l	98
11) Tert butyl alcohol	5.536	59	28940	76.789	ug/l	96
12) 1,1-Dichloroethene	4.353	96	30242	19.526	ug/l	92
13) Acrolein	4.194	56	26768	81.530	ug/l	97
14) Allyl chloride	5.036	41	36755	18.346	ug/l	94
15) Acrylonitrile	5.730	53	81441	94.188	ug/l	100
16) Acetone	4.442	43	81191	96.955	ug/l	98
17) Carbon Disulfide	4.724	76	69558	17.281	ug/l	97
18) Methyl Acetate	5.036	43	44110	20.120	ug/l	97
19) Methyl tert-butyl Ether	5.806	73	108812	19.832	ug/l	100
20) Methylene Chloride	5.294	84	36799	20.927	ug/l	87
21) trans-1,2-Dichloroethene	5.800	96	32061	18.164	ug/l	93
22) Diisopropyl ether	6.683	45	97238	21.269	ug/l	95
23) Vinyl Acetate	6.618	43	357714m	98.349	ug/l	
24) 1,1-Dichloroethane	6.583	63	59081	20.133	ug/l	99
25) 2-Butanone	7.488	43	109556	93.707	ug/l	97
26) 2,2-Dichloropropane	7.500	77	50087	18.365	ug/l	96
27) cis-1,2-Dichloroethene	7.494	96	42802	21.445	ug/l	92
28) Bromochloromethane	7.824	49	25359	21.289	ug/l	86
29) Tetrahydrofuran	7.841	42	66219	93.116	ug/l	99
30) Chloroform	7.971	83	67707	20.091	ug/l	91
31) Cyclohexane	8.259	56	52027	18.240	ug/l	96
32) 1,1,1-Trichloroethane	8.177	97	61258	20.020	ug/l	94
36) 1,1-Dichloropropene	8.377	75	45547	19.565	ug/l	97
37) Ethyl Acetate	7.571	43	41666	19.877	ug/l	99
38) Carbon Tetrachloride	8.371	117	55209	21.589	ug/l	95
39) Methylcyclohexane	9.606	83	56098	19.254	ug/l	98
40) Benzene	8.612	78	144698	20.301	ug/l	97

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN120522\
 Data File : VN075558.D
 Acq On : 05 Dec 2022 14:55
 Operator : JC\MD
 Sample : VN1205WBSD01
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VN1205WBSD01

Manual Integrations
 APPROVED

Reviewed By :Krupa Patel 12/06/2022
 Supervised By :Mahesh Dadoda 12/06/2022

Quant Time: Dec 06 03:24:44 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N112122W.M
 Quant Title : SW846 8260
 QLast Update : Wed Nov 23 00:18:03 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
41) Methacrylonitrile	7.788	41	21774	20.318	ug/l	99
42) 1,2-Dichloroethane	8.677	62	51074	21.396	ug/l	98
43) Isopropyl Acetate	8.694	43	66829	19.862	ug/l	96
44) Trichloroethene	9.359	130	40250	20.638	ug/l	88
45) 1,2-Dichloropropane	9.624	63	35752	21.049	ug/l	97
46) Dibromomethane	9.712	93	26418	20.563	ug/l	90
47) Bromodichloromethane	9.888	83	55155	22.169	ug/l	96
48) Methyl methacrylate	9.682	41	30699	19.732	ug/l	97
49) 1,4-Dioxane	9.700	88	15796	357.561	ug/l	97
51) 4-Methyl-2-Pentanone	10.447	43	213961	99.843	ug/l	97
52) Toluene	10.635	92	95079	20.701	ug/l	99
53) t-1,3-Dichloropropene	10.841	75	52520	21.027	ug/l	96
54) cis-1,3-Dichloropropene	10.318	75	58271	21.481	ug/l	96
55) 1,1,2-Trichloroethane	11.023	97	38649	20.660	ug/l	97
56) Ethyl methacrylate	10.876	69	52573	20.810	ug/l	98
57) 1,3-Dichloropropane	11.165	76	63536	20.830	ug/l	99
58) 2-Chloroethyl Vinyl ether	10.165	63	91122	100.008	ug/l	97
59) 2-Hexanone	11.200	43	161598	100.423	ug/l	99
60) Dibromochloromethane	11.359	129	45759	23.830	ug/l	99
61) 1,2-Dibromoethane	11.470	107	39070	20.122	ug/l	99
64) Tetrachloroethene	11.106	164	35145	18.672	ug/l	97
65) Chlorobenzene	11.894	112	104956	20.619	ug/l	100
66) 1,1,1,2-Tetrachloroethane	11.965	131	41050	21.608	ug/l	100
67) Ethyl Benzene	11.965	91	180449	20.779	ug/l	99
68) m/p-Xylenes	12.070	106	147054	41.699	ug/l	99
69) o-Xylene	12.400	106	74105	21.149	ug/l	95
70) Styrene	12.418	104	120009	21.711	ug/l	97
71) Bromoform	12.582	173	33484	23.158	ug/l #	98
73) Isopropylbenzene	12.700	105	188042	20.328	ug/l	99
74) N-amyl acetate	12.500	43	55535	19.527	ug/l	98
75) 1,1,2,2-Tetrachloroethane	12.941	83	57856	20.514	ug/l	100
76) 1,2,3-Trichloropropane	12.994	75	48220m	20.142	ug/l	
77) Bromobenzene	12.982	156	47578	20.833	ug/l	84
78) n-propylbenzene	13.041	91	212278	20.562	ug/l	99
79) 2-Chlorotoluene	13.129	91	128287	20.624	ug/l	97
80) 1,3,5-Trimethylbenzene	13.176	105	161221	20.977	ug/l	99
81) trans-1,4-Dichloro-2-b...	12.741	75	18561	22.940	ug/l	86
82) 4-Chlorotoluene	13.129	91	128287	20.624	ug/l	97
83) tert-Butylbenzene	13.441	119	143196	20.435	ug/l	96
84) 1,2,4-Trimethylbenzene	13.488	105	165255	21.437	ug/l	99
85) sec-Butylbenzene	13.617	105	198139	20.839	ug/l	99
86) p-Isopropyltoluene	13.735	119	168958	21.121	ug/l	99
87) 1,3-Dichlorobenzene	13.735	146	87445	20.209	ug/l	97
88) 1,4-Dichlorobenzene	13.817	146	87432	19.891	ug/l	99
89) n-Butylbenzene	14.059	91	132558	19.620	ug/l	98
90) Hexachloroethane	14.335	117	30050	21.775	ug/l	81
91) 1,2-Dichlorobenzene	14.106	146	88485	20.979	ug/l	97
92) 1,2-Dibromo-3-Chloropr...	14.723	75	10901	20.405	ug/l	84
93) 1,2,4-Trichlorobenzene	15.400	180	45024	20.830	ug/l	98
94) Hexachlorobutadiene	15.500	225	23896	21.225	ug/l	97
95) Naphthalene	15.647	128	130927	19.254	ug/l	100
96) 1,2,3-Trichlorobenzene	15.847	180	43705	20.573	ug/l	98

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN120522\
Data File : VN075558.D
Acq On : 05 Dec 2022 14:55
Operator : JC\MD
Sample : VN1205WBSD01
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VN1205WBSD01

Manual Integrations
APPROVED

Reviewed By :Krupa Patel 12/06/2022
Supervised By :Mahesh Dadoda 12/06/2022

Quant Time: Dec 06 03:24:44 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N112122W.M
Quant Title : SW846 8260
QLast Update : Wed Nov 23 00:18:03 2022
Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
----------	------	------	----------	------	-------	----------

(#) = qualifier out of range (m) = manual integration (+) = signals summed


Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN120522\
 Data File : VN075558.D
 Acq On : 05 Dec 2022 14:55
 Operator : JC\MD
 Sample : VN1205WBSD01
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 8 Sample Multiplier: 1


Instrument :
 MSVOA_N
 ClientSampleId :
 VN1205WBSD01

Manual Integrations APPROVED

Reviewed By :Krupa Patel 12/06/2022
 Supervised By :Mahesh Dadoda 12/06/2022


Quant Time: Dec 06 03:24:44 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N112122W.M
 Quant Title : SW846 8260
 QLast Update : Wed Nov 23 00:18:03 2022
 Response via : Initial Calibration





#1
Pentafluorobenzene
Concen: 50.000 ug/l
RT: 8.230 min Scan# 1050
Delta R.T. -0.000 min
Lab File: VN075558.D
Acq: 05 Dec 2022 14:55




Instrument :
MSVOA_N
ClientSampleId :
VN1205WBSD01



Tgt Ion:168 Resp: 16725
Ion Ratio Lower Upper
168 100
99 55.9 51.6 77.4




Manual Integrations
APPROVED


Reviewed By :Krupa Patel 12/06/2022
Supervised By :Mahesh Dadoda 12/06/2022



#2
Dichlorodifluoromethane
Concen: 18.173 ug/l
RT: 2.136 min Scan# 14
Delta R.T. -0.000 min
Lab File: VN075558.D
Acq: 05 Dec 2022 14:55

Tgt Ion: 85 Resp: 23241
Ion Ratio Lower Upper
85 100
87 32.0 16.7 50.0









#3
Chloromethane
Concen: 19.482 ug/l
RT: 2.371 min Scan# 54
Delta R.T. 0.000 min
Lab File: VN075558.D
Acq: 05 Dec 2022 14:55

Instrument :
MSVOA_N
ClientSampleId :
VN1205WBSD01

Tgt Ion: 50 Resp: 3100
Ion Ratio Lower Upper
50 100
52 32.9 25.8 38.6




Manual Integrations
APPROVED


Reviewed By :Krupa Patel 12/06/2022
Supervised By :Mahesh Dadoda 12/06/2022



#4
Vinyl Chloride
Concen: 19.188 ug/l
RT: 2.524 min Scan# 80
Delta R.T. 0.000 min
Lab File: VN075558.D
Acq: 05 Dec 2022 14:55

Tgt Ion: 62 Resp: 38482
Ion Ratio Lower Upper
62 100
64 33.8 24.0 36.0









#5
Bromomethane
Concen: 19.226 ug/l
RT: 2.948 min Scan# 11
Delta R.T. 0.030 min
Lab File: VN075558.D
Acq: 05 Dec 2022 14:55

Instrument :
MSVOA_N
ClientSampleId :
VN1205WBSD01

Tgt Ion: 94 Resp: 3390
Ion Ratio Lower Upper
94 100
96 86.1 71.8 107.6




Manual Integrations
APPROVED


Reviewed By :Krupa Patel 12/06/2022
Supervised By :Mahesh Dadoda 12/06/2022



#6
Chloroethane
Concen: 21.113 ug/l
RT: 3.124 min Scan# 182
Delta R.T. 0.017 min
Lab File: VN075558.D
Acq: 05 Dec 2022 14:55

Tgt Ion: 64 Resp: 32085
Ion Ratio Lower Upper
64 100
66 32.6 23.8 35.8









#7
Trichlorofluoromethane
Concen: 18.744 ug/l
RT: 3.506 min Scan# 246
Delta R.T. 0.005 min
Lab File: VN075558.D
Acq: 05 Dec 2022 14:55

Instrument :
MSVOA_N
ClientSampleId :
VN1205WBSD01

Tgt Ion:101 Resp: 5370
Ion Ratio Lower Upper
101 100
103 66.3 50.5 75.7




Manual Integrations
APPROVED


Reviewed By :Krupa Patel 12/06/2022
Supervised By :Mahesh Dadoda 12/06/2022



#8
Diethyl Ether
Concen: 21.245 ug/l
RT: 3.971 min Scan# 326
Delta R.T. 0.006 min
Lab File: VN075558.D
Acq: 05 Dec 2022 14:55


Tgt Ion: 74 Resp: 21334
Ion Ratio Lower Upper
74 100
45 77.6 39.3 117.9





#9
1,1,2-Trichlorotrifluoroethane
Concen: 19.884 ug/l
RT: 4.383 min Scan# 396
Delta R.T. 0.006 min
Lab File: VN075558.D
Acq: 05 Dec 2022 14:55




Instrument :
MSVOA_N
ClientSampleId :
VN1205WBSD01



Tgt Ion:101 Resp: 31450
Ion Ratio Lower Upper
101 100
85 44.7 37.6 56.4
151 79.7 58.3 87.5




Manual Integrations
APPROVED

Reviewed By :Krupa Patel 12/06/2022
Supervised By :Mahesh Dadoda 12/06/2022



#10
Methyl Iodide
Concen: 19.075 ug/l
RT: 4.600 min Scan# 433
Delta R.T. 0.005 min
Lab File: VN075558.D
Acq: 05 Dec 2022 14:55


Tgt Ion:142 Resp: 48889
Ion Ratio Lower Upper
142 100
127 45.6 38.0 57.0
141 15.0 12.2 18.4





#11
Tert butyl alcohol
Concen: 76.789 ug/l
RT: 5.536 min Scan# 593
Delta R.T. -0.006 min
Lab File: VN075558.D
Acq: 05 Dec 2022 14:55




Instrument :
MSVOA_N
ClientSampleId :
VN1205WBSD01




Tgt Ion: 59 Resp: 28940
Ion Ratio Lower Upper
59 100
57 9.1 8.5 12.7

Manual Integrations
APPROVED



Reviewed By :Krupa Patel 12/06/2022
Supervised By :Mahesh Dadoda 12/06/2022




#12
1,1-Dichloroethene
Concen: 19.526 ug/l
RT: 4.353 min Scan# 391
Delta R.T. 0.011 min
Lab File: VN075558.D
Acq: 05 Dec 2022 14:55



Tgt Ion: 96 Resp: 30242
Ion Ratio Lower Upper
96 100
61 147.1 126.2 189.4
98 59.7 51.5 77.3









#13
Acrolein
Concen: 81.530 ug/l
RT: 4.194 min Scan# 364
Delta R.T. 0.011 min
Lab File: VN075558.D
Acq: 05 Dec 2022 14:55

Instrument :
MSVOA_N
ClientSampleId :
VN1205WBSD01

Tgt Ion: 56 Resp: 26768
Ion Ratio Lower Upper
56 100
55 71.2 55.2 82.8




Manual Integrations
APPROVED


Reviewed By :Krupa Patel 12/06/2022
Supervised By :Mahesh Dadoda 12/06/2022



#14
Allyl chloride
Concen: 18.346 ug/l
RT: 5.036 min Scan# 507
Delta R.T. -0.000 min
Lab File: VN075558.D
Acq: 05 Dec 2022 14:55


Tgt Ion: 41 Resp: 36755
Ion Ratio Lower Upper
41 100
39 80.6 61.0 91.4
76 46.6 33.4 50.0





#15
Acrylonitrile
Concen: 94.188 ug/l
RT: 5.730 min Scan# 61
Delta R.T. 0.006 min
Lab File: VN075558.D
Acq: 05 Dec 2022 14:55




Instrument :
MSVOA_N
ClientSampleId :
VN1205WBSD01



Tgt Ion: 53 Resp: 8144
Ion Ratio Lower Upper
53 100
52 82.6 66.4 99.6
51 37.1 29.6 44.4




Manual Integrations
APPROVED


Reviewed By :Krupa Patel 12/06/2022
Supervised By :Mahesh Dadoda 12/06/2022



#16
Acetone
Concen: 96.955 ug/l
RT: 4.442 min Scan# 406
Delta R.T. 0.006 min
Lab File: VN075558.D
Acq: 05 Dec 2022 14:55

Tgt Ion: 43 Resp: 81191
Ion Ratio Lower Upper
43 100
58 33.1 25.5 38.3









#17
Carbon Disulfide
Concen: 17.281 ug/l
RT: 4.724 min Scan# 453
Delta R.T. 0.006 min
Lab File: VN075558.D
Acq: 05 Dec 2022 14:55

Instrument :
MSVOA_N
ClientSampleId :
VN1205WBSD01

Tgt Ion: 76 Resp: 69558
Ion Ratio Lower Upper
76 100
78 8.1 7.4 11.2




Manual Integrations
APPROVED


Reviewed By :Krupa Patel 12/06/2022
Supervised By :Mahesh Dadoda 12/06/2022



#18
Methyl Acetate
Concen: 20.120 ug/l
RT: 5.036 min Scan# 507
Delta R.T. -0.000 min
Lab File: VN075558.D
Acq: 05 Dec 2022 14:55

Tgt Ion: 43 Resp: 44110
Ion Ratio Lower Upper
43 100
74 29.2 22.0 33.0









#19
Methyl tert-butyl Ether
Concen: 19.832 ug/l
RT: 5.806 min Scan# 61
Delta R.T. 0.000 min
Lab File: VN075558.D
Acq: 05 Dec 2022 14:55

Instrument :
MSVOA_N
ClientSampleId :
VN1205WBSD01

Tgt Ion: 73 Resp: 10881
Ion Ratio Lower Upper
73 100
57 20.5 16.3 24.5




Manual Integrations
APPROVED


Reviewed By :Krupa Patel 12/06/2022
Supervised By :Mahesh Dadoda 12/06/2022



#20
Methylene Chloride
Concen: 20.927 ug/l
RT: 5.294 min Scan# 551
Delta R.T. 0.005 min
Lab File: VN075558.D
Acq: 05 Dec 2022 14:55

Tgt Ion: 84 Resp: 36799
Ion Ratio Lower Upper
84 100
49 100.4 94.9 142.3
51 28.3 26.8 40.2
86 59.5 53.7 80.5









#21
trans-1,2-Dichloroethene
Concen: 18.164 ug/l
RT: 5.800 min Scan# 61
Delta R.T. 0.005 min
Lab File: VN075558.D
Acq: 05 Dec 2022 14:55

Instrument :
MSVOA_N
ClientSampleId :
VN1205WBSD01

Tgt Ion: 96 Resp: 3206
Ion Ratio Lower Upper
96 100
61 116.8 99.9 149.9
98 58.7 51.5 77.3




Manual Integrations
APPROVED

Reviewed By :Krupa Patel 12/06/2022
Supervised By :Mahesh Dadoda 12/06/2022



#22
Diisopropyl ether
Concen: 21.269 ug/l
RT: 6.683 min Scan# 787
Delta R.T. 0.006 min
Lab File: VN075558.D
Acq: 05 Dec 2022 14:55

Tgt Ion: 45 Resp: 97238
Ion Ratio Lower Upper
45 100
43 57.8 42.8 64.2
87 38.6 28.4 42.6
59 13.5 10.4 15.6









#23
 Vinyl Acetate
 Concen: 98.349 ug/l m
 RT: 6.618 min Scan# 714
 Delta R.T. 0.006 min
 Lab File: VN075558.D
 Acq: 05 Dec 2022 14:55

Instrument :
 MSVOA_N
 ClientSampleId :
 VN1205WBSD01

Tgt Ion: 43 Resp: 357714
 Ion Ratio Lower Upper
 43 100
 86 14.2 10.7 16.1




Manual Integrations
 APPROVED


Reviewed By :Krupa Patel 12/06/2022
 Supervised By :Mahesh Dadoda 12/06/2022



#24
 1,1-Dichloroethane
 Concen: 20.133 ug/l
 RT: 6.583 min Scan# 770
 Delta R.T. 0.006 min
 Lab File: VN075558.D
 Acq: 05 Dec 2022 14:55


Tgt Ion: 63 Resp: 59081
 Ion Ratio Lower Upper
 63 100
 98 7.4 3.6 10.8
 100 5.5 2.4 7.2





#25
2-Butanone
Concen: 93.707 ug/l
RT: 7.488 min Scan# 912
Delta R.T. -0.001 min
Lab File: VN075558.D
Acq: 05 Dec 2022 14:55




Instrument :
MSVOA_N
ClientSampleId :
VN1205WBSD01




Tgt Ion: 43 Resp: 109556
Ion Ratio Lower Upper
43 100
72 30.1 25.5 38.3

Manual Integrations
APPROVED



Reviewed By :Krupa Patel 12/06/2022
Supervised By :Mahesh Dadoda 12/06/2022




#26
2,2-Dichloropropane
Concen: 18.365 ug/l
RT: 7.500 min Scan# 926
Delta R.T. 0.006 min
Lab File: VN075558.D
Acq: 05 Dec 2022 14:55



Tgt Ion: 77 Resp: 50087
Ion Ratio Lower Upper
77 100
97 24.5 11.2 33.6






#27
 cis-1,2-Dichloroethene
 Concen: 21.445 ug/l
 RT: 7.494 min Scan# 91
 Delta R.T. 0.000 min
 Lab File: VN075558.D
 Acq: 05 Dec 2022 14:55

Instrument :

MSVOA_N

ClientSampleId :

VN1205WBSD01






Tgt Ion: 96 Resp: 4280
 Ion Ratio Lower Upper
 96 100
 61 125.5 0.0 278.8
 98 63.9 0.0 128.4

Manual Integrations

APPROVED




Reviewed By :Krupa Patel 12/06/2022


Supervised By :Mahesh Dadoda 12/06/2022



#28
 Bromochloromethane
 Concen: 21.289 ug/l
 RT: 7.824 min Scan# 981
 Delta R.T. 0.006 min
 Lab File: VN075558.D
 Acq: 05 Dec 2022 14:55

Tgt Ion: 49 Resp: 25359
 Ion Ratio Lower Upper
 49 100
 129 1.9 0.0 4.6
 130 95.7 66.2 99.4








#29
Tetrahydrofuran
Concen: 93.116 ug/l
RT: 7.841 min Scan# 984
Delta R.T. 0.000 min
Lab File: VN075558.D
Acq: 05 Dec 2022 14:55

Instrument :
MSVOA_N
ClientSampleId :
VN1205WBSD01

Tgt Ion: 42 Resp: 66219
Ion Ratio Lower Upper
42 100
72 53.7 43.0 64.6
71 51.0 39.9 59.9




Manual Integrations
APPROVED


Reviewed By :Krupa Patel 12/06/2022
Supervised By :Mahesh Dadoda 12/06/2022



#30
Chloroform
Concen: 20.091 ug/l
RT: 7.971 min Scan# 1006
Delta R.T. -0.000 min
Lab File: VN075558.D
Acq: 05 Dec 2022 14:55


Tgt Ion: 83 Resp: 67707
Ion Ratio Lower Upper
83 100
85 57.0 51.4 77.0





#31
Cyclohexane
Concen: 18.240 ug/l
RT: 8.259 min Scan# 1055
Delta R.T. -0.000 min
Lab File: VN075558.D
Acq: 05 Dec 2022 14:55




Instrument :
MSVOA_N
ClientSampleId :
VN1205WBSD01




Tgt Ion: 56 Resp: 5202
Ion Ratio Lower Upper
56 100
69 32.1 25.2 37.8
84 92.2 77.8 116.8

Manual Integrations
APPROVED



Reviewed By :Krupa Patel 12/06/2022
Supervised By :Mahesh Dadoda 12/06/2022



#32
1,1,1-Trichloroethane
Concen: 20.020 ug/l
RT: 8.177 min Scan# 1041
Delta R.T. 0.006 min
Lab File: VN075558.D
Acq: 05 Dec 2022 14:55



Tgt Ion: 97 Resp: 61258
Ion Ratio Lower Upper
97 100
99 60.4 50.2 75.2
61 39.4 37.0 55.6





#33

1,2-Dichloroethane-d4

Concen: 52.415 ug/l

RT: 8.582 min Scan# 1110

Delta R.T. -0.001 min

Lab File: VN075558.D

Acq: 05 Dec 2022 14:55

Instrument :

MSVOA_N

ClientSampleId :

VN1205WBSD01

Tgt Ion: 65 Resp: 4705

Ion Ratio Lower Upper

65 100





67 51.0 0.0 102.0

Manual Integrations

APPROVED

Reviewed By :Krupa Patel 12/06/2022

Supervised By :Mahesh Dadoda 12/06/2022



#34

1,4-Difluorobenzene

Concen: 50.000 ug/l

RT: 9.106 min Scan# 1199

Delta R.T. -0.000 min

Lab File: VN075558.D

Acq: 05 Dec 2022 14:55




Tgt Ion:114 Resp: 271743

Ion Ratio Lower Upper

114 100

63 17.6 0.0 39.6

88 14.7 0.0 29.0





#35

Dibromofluoromethane

Concen: 56.888 ug/l

RT: 8.177 min Scan# 1040

Delta R.T. 0.006 min

Lab File: VN075558.D

Acq: 05 Dec 2022 14:55

Instrument :

MSVOA_N

ClientSampleId :

VN1205WBSD01

Tgt Ion:113 Resp: 60728





Ion	Ratio	Lower	Upper
113	100		
111	101.4	80.5	120.7
192	19.3	13.2	19.8

Manual Integrations

APPROVED

Reviewed By :Krupa Patel 12/06/2022

Supervised By :Mahesh Dadoda 12/06/2022



#36

1,1-Dichloropropene

Concen: 19.565 ug/l

RT: 8.377 min Scan# 1075




Delta R.T. 0.006 min


Lab File: VN075558.D

Acq: 05 Dec 2022 14:55

Tgt Ion: 75 Resp: 45547


Ion	Ratio	Lower	Upper
75	100		
110	38.1	17.8	53.3
77	31.5	24.6	37.0





#37
Ethyl Acetate
Concen: 19.877 ug/l
RT: 7.571 min Scan# 91
Delta R.T. 0.006 min
Lab File: VN075558.D
Acq: 05 Dec 2022 14:55




Instrument :
MSVOA_N
ClientSampleId :
VN1205WBSD01




Tgt Ion: 43 Resp: 4166
Ion Ratio Lower Upper
43 100
61 15.0 12.2 18.4
70 11.9 9.4 14.2

Manual Integrations
APPROVED



Reviewed By :Krupa Patel 12/06/2022
Supervised By :Mahesh Dadoda 12/06/2022



#38
Carbon Tetrachloride
Concen: 21.589 ug/l
RT: 8.371 min Scan# 1074
Delta R.T. 0.006 min
Lab File: VN075558.D
Acq: 05 Dec 2022 14:55



Tgt Ion:117 Resp: 55209
Ion Ratio Lower Upper
117 100
119 90.8 76.7 115.1
121 31.5 23.9 35.9





#39

Methylcyclohexane

Concen: 19.254 ug/l

RT: 9.606 min Scan# 1115

Delta R.T. 0.006 min

Lab File: VN075558.D

Acq: 05 Dec 2022 14:55

Instrument :

MSVOA_N

ClientSampleId :

VN1205WBSD01





Tgt Ion: 83 Resp: 56098
 Ion Ratio Lower Upper
 83 100
 55 68.6 53.6 80.4
 98 53.1 41.1 61.7

Manual Integrations

APPROVED

Reviewed By :Krupa Patel 12/06/2022

Supervised By :Mahesh Dadoda 12/06/2022



#40

Benzene

Concen: 20.301 ug/l




RT: 8.612 min Scan# 1115

Delta R.T. 0.006 min

Lab File: VN075558.D

Acq: 05 Dec 2022 14:55

Tgt Ion: 78 Resp: 144698
 Ion Ratio Lower Upper
 78 100
 77 22.9 19.4 29.0





#41
Methacrylonitrile
Concen: 20.318 ug/l
RT: 7.788 min Scan# 91
Delta R.T. 0.005 min
Lab File: VN075558.D
Acq: 05 Dec 2022 14:55





Instrument :
MSVOA_N
ClientSampleId :
VN1205WBSD01

Tgt Ion: 41 Resp: 21774

Ion	Ratio	Lower	Upper
41	100		
39	56.1	44.4	66.6
67	89.3	69.8	104.6
52	33.9	27.3	40.9

Manual Integrations
APPROVED




Reviewed By :Krupa Patel 12/06/2022
Supervised By :Mahesh Dadoda 12/06/2022




#42
1,2-Dichloroethane
Concen: 21.396 ug/l
RT: 8.677 min Scan# 1126
Delta R.T. 0.006 min
Lab File: VN075558.D
Acq: 05 Dec 2022 14:55

Tgt Ion: 62 Resp: 51074

Ion	Ratio	Lower	Upper
62	100		
98	9.8	0.0	18.2





#43

Isopropyl Acetate

Concen: 19.862 ug/l

RT: 8.694 min Scan# 1129

Delta R.T. 0.006 min

Lab File: VN075558.D

Acq: 05 Dec 2022 14:55

Instrument :

MSVOA_N

ClientSampleId :

VN1205WBSD01

Tgt Ion: 43 Resp: 66829





Ion	Ratio	Lower	Upper
43	100		
61	30.7	22.4	33.6
87	16.5	12.5	18.7

Manual Integrations

APPROVED

Reviewed By :Krupa Patel 12/06/2022

Supervised By :Mahesh Dadoda 12/06/2022



#44

Trichloroethene

Concen: 20.638 ug/l

RT: 9.359 min Scan# 1242




Delta R.T. 0.006 min


Lab File: VN075558.D

Acq: 05 Dec 2022 14:55

Tgt Ion:130 Resp: 40250

Ion	Ratio	Lower	Upper
130	100		
95	87.4	0.0	198.4





#45

1,2-Dichloropropane

Concen: 21.049 ug/l

RT: 9.624 min Scan# 1287

Delta R.T. -0.000 min

Lab File: VN075558.D

Acq: 05 Dec 2022 14:55

Instrument :

MSVOA_N

ClientSampleId :

VN1205WBSD01

Tgt Ion: 63 Resp: 3575

Ion Ratio Lower Upper

63 100





65 34.3 26.1 39.1

Manual Integrations

APPROVED

Reviewed By :Krupa Patel 12/06/2022

Supervised By :Mahesh Dadoda 12/06/2022



#46

Dibromomethane

Concen: 20.563 ug/l

RT: 9.712 min Scan# 1302

Delta R.T. -0.000 min

Lab File: VN075558.D

Acq: 05 Dec 2022 14:55




Tgt Ion: 93 Resp: 26418

Ion Ratio Lower Upper

93 100

95 83.4 66.6 99.8


174 105.3 70.3 105.5





#47
Bromodichloromethane
Concen: 22.169 ug/l
RT: 9.888 min Scan# 1332
Delta R.T. 0.000 min
Lab File: VN075558.D
Acq: 05 Dec 2022 14:55




Instrument :
MSVOA_N
ClientSampleId :
VN1205WBSD01



Tgt Ion: 83 Resp: 5515
Ion Ratio Lower Upper
83 100
85 62.0 52.1 78.1
127 8.0 6.6 10.0




Manual Integrations
APPROVED


Reviewed By :Krupa Patel 12/06/2022
Supervised By :Mahesh Dadoda 12/06/2022



#48
Methyl methacrylate
Concen: 19.732 ug/l
RT: 9.682 min Scan# 1297
Delta R.T. -0.001 min
Lab File: VN075558.D
Acq: 05 Dec 2022 14:55


Tgt Ion: 41 Resp: 30699
Ion Ratio Lower Upper
41 100
69 102.2 80.2 120.2
39 66.1 56.2 84.2





#49
1,4-Dioxane
Concen: 357.561 ug/l
RT: 9.700 min Scan# 11
Delta R.T. 0.000 min
Lab File: VN075558.D
Acq: 05 Dec 2022 14:55




Instrument :
MSVOA_N
ClientSampleId :
VN1205WBSD01




Tgt Ion: 88 Resp: 15790
Ion Ratio Lower Upper
88 100
43 24.1 17.5 26.3
58 60.7 49.8 74.8

Manual Integrations
APPROVED



Reviewed By :Krupa Patel 12/06/2022
Supervised By :Mahesh Dadoda 12/06/2022




#50
Toluene-d8
Concen: 52.481 ug/l
RT: 10.571 min Scan# 1448
Delta R.T. 0.006 min
Lab File: VN075558.D
Acq: 05 Dec 2022 14:55



Tgt Ion: 98 Resp: 126181
Ion Ratio Lower Upper
98 100
100 66.6 53.0 79.6









#51
4-Methyl-2-Pentanone
Concen: 99.843 ug/l
RT: 10.447 min Scan# 1427
Delta R.T. 0.000 min
Lab File: VN075558.D
Acq: 05 Dec 2022 14:55

Instrument :
MSVOA_N
ClientSampleId :
VN1205WBSD01

Tgt Ion: 43 Resp: 21396
Ion Ratio Lower Upper
43 100
58 43.8 33.6 50.4




Manual Integrations
APPROVED


Reviewed By :Krupa Patel 12/06/2022
Supervised By :Mahesh Dadoda 12/06/2022



#52
Toluene
Concen: 20.701 ug/l
RT: 10.635 min Scan# 1459
Delta R.T. 0.005 min
Lab File: VN075558.D
Acq: 05 Dec 2022 14:55

Tgt Ion: 92 Resp: 95079
Ion Ratio Lower Upper
92 100
91 171.2 135.4 203.2









#53
 t-1,3-Dichloropropene
 Concen: 21.027 ug/l
 RT: 10.841 min Scan# 1493
 Delta R.T. 0.006 min
 Lab File: VN075558.D
 Acq: 05 Dec 2022 14:55

Instrument :
 MSVOA_N
 ClientSampleId :
 VN1205WBSD01

Tgt Ion: 75 Resp: 52520
 Ion Ratio Lower Upper
 75 100
 77 29.6 25.4 38.2




Manual Integrations
 APPROVED


Reviewed By :Krupa Patel 12/06/2022
 Supervised By :Mahesh Dadoda 12/06/2022



#54
 cis-1,3-Dichloropropene
 Concen: 21.481 ug/l
 RT: 10.318 min Scan# 1405
 Delta R.T. 0.006 min
 Lab File: VN075558.D
 Acq: 05 Dec 2022 14:55


Tgt Ion: 75 Resp: 58271
 Ion Ratio Lower Upper
 75 100
 77 32.8 23.7 35.5
 39 42.6 35.8 53.6





#55
1,1,2-Trichloroethane
Concen: 20.660 ug/l
RT: 11.023 min Scan# 1100
Delta R.T. 0.005 min
Lab File: VN075558.D
Acq: 05 Dec 2022 14:55

Instrument :
MSVOA_N
ClientSampleId :
VN1205WBSD01






Tgt Ion: 97 Resp: 38649


Ion	Ratio	Lower	Upper
97	100		
83	79.8	65.9	98.9
85	49.0	40.4	60.6
99	61.9	46.4	69.6

Manual Integrations
APPROVED

Reviewed By :Krupa Patel 12/06/2022
Supervised By :Mahesh Dadoda 12/06/2022






#56
Ethyl methacrylate
Concen: 20.810 ug/l
RT: 10.876 min Scan# 1500
Delta R.T. -0.001 min
Lab File: VN075558.D
Acq: 05 Dec 2022 14:55



Tgt Ion: 69 Resp: 52573

Ion	Ratio	Lower	Upper
69	100		
41	55.5	45.6	68.4
39	34.7	28.5	42.7





#57

1,3-Dichloropropane

Concen: 20.830 ug/l

RT: 11.165 min Scan# 1

Delta R.T. -0.000 min

Lab File: VN075558.D

Acq: 05 Dec 2022 14:55

Instrument :

MSVOA_N

ClientSampleId :

VN1205WBSD01

Tgt Ion: 76 Resp: 63530

Ion Ratio Lower Upper

76 100





78 32.7 25.8 38.6

Manual Integrations

APPROVED

Reviewed By :Krupa Patel 12/06/2022

Supervised By :Mahesh Dadoda 12/06/2022



#58

2-Chloroethyl Vinyl ether

Concen: 100.008 ug/l

RT: 10.165 min Scan# 1379

Delta R.T. 0.006 min

Lab File: VN075558.D




Acq: 05 Dec 2022 14:55


Tgt Ion: 63 Resp: 91122

Ion Ratio Lower Upper

63 100


106 32.8 24.8 37.2





#59
2-Hexanone
Concen: 100.423 ug/l
RT: 11.200 min Scan# 1555
Delta R.T. 0.006 min
Lab File: VN075558.D
Acq: 05 Dec 2022 14:55




Instrument :
MSVOA_N
ClientSampleId :
VN1205WBSD01



Tgt Ion: 43 Resp: 161598
Ion Ratio Lower Upper
43 100
58 59.9 29.6 88.8




Manual Integrations
APPROVED

Reviewed By :Krupa Patel 12/06/2022
Supervised By :Mahesh Dadoda 12/06/2022



#60
Dibromochloromethane
Concen: 23.830 ug/l
RT: 11.359 min Scan# 1582
Delta R.T. -0.000 min
Lab File: VN075558.D
Acq: 05 Dec 2022 14:55


Tgt Ion:129 Resp: 45759
Ion Ratio Lower Upper
129 100
127 76.6 38.6 115.8





#61
1,2-Dibromoethane
Concen: 20.122 ug/l
RT: 11.470 min Scan# 1601
Delta R.T. -0.001 min
Lab File: VN075558.D
Acq: 05 Dec 2022 14:55




Instrument :
MSVOA_N
ClientSampleId :
VN1205WBSD01




Tgt Ion:107 Resp: 39070
Ion Ratio Lower Upper
107 100
109 95.3 75.6 113.4

Manual Integrations
APPROVED



Reviewed By :Krupa Patel 12/06/2022
Supervised By :Mahesh Dadoda 12/06/2022




#62
4-Bromofluorobenzene
Concen: 54.802 ug/l
RT: 12.853 min Scan# 1836
Delta R.T. 0.006 min
Lab File: VN075558.D
Acq: 05 Dec 2022 14:55




Tgt Ion: 95 Resp: 97253
Ion Ratio Lower Upper
95 100
174 87.6 0.0 147.4
176 81.8 0.0 142.2





#63
Chlorobenzene-d5
Concen: 50.000 ug/l
RT: 11.870 min Scan# 1669
Delta R.T. 0.005 min
Lab File: VN075558.D
Acq: 05 Dec 2022 14:55




Instrument :
MSVOA_N
ClientSampleId :
VN1205WBSD01



Tgt Ion:117 Resp: 24648
Ion Ratio Lower Upper
117 100
82 51.6 43.1 64.7
119 32.2 25.5 38.3




Manual Integrations
APPROVED


Reviewed By :Krupa Patel 12/06/2022
Supervised By :Mahesh Dadoda 12/06/2022



#64
Tetrachloroethene
Concen: 18.672 ug/l
RT: 11.106 min Scan# 1539
Delta R.T. -0.000 min
Lab File: VN075558.D
Acq: 05 Dec 2022 14:55


Tgt Ion:164 Resp: 35145
Ion Ratio Lower Upper
164 100
166 129.6 106.5 159.7
129 103.1 81.9 122.9
131 96.4 81.4 122.2





#65
Chlorobenzene
Concen: 20.619 ug/l
RT: 11.894 min Scan# 1673
Delta R.T. 0.000 min
Lab File: VN075558.D
Acq: 05 Dec 2022 14:55




Instrument :
MSVOA_N
ClientSampleId :
VN1205WBSD01



Tgt Ion:112 Resp: 104950
Ion Ratio Lower Upper
112 100
114 31.9 25.4 38.2




Manual Integrations
APPROVED


Reviewed By :Krupa Patel 12/06/2022
Supervised By :Mahesh Dadoda 12/06/2022



#66
1,1,1,2-Tetrachloroethane
Concen: 21.608 ug/l
RT: 11.965 min Scan# 1685
Delta R.T. -0.000 min
Lab File: VN075558.D
Acq: 05 Dec 2022 14:55


Tgt Ion:131 Resp: 41050
Ion Ratio Lower Upper
131 100
133 95.5 47.8 143.4
119 65.3 33.0 99.0





#67
Ethyl Benzene
Concen: 20.779 ug/l
RT: 11.965 min Scan# 1685
Delta R.T. -0.000 min
Lab File: VN075558.D
Acq: 05 Dec 2022 14:55




Instrument :
MSVOA_N
ClientSampleId :
VN1205WBSD01




Tgt Ion: 91 Resp: 180449
Ion Ratio Lower Upper
91 100
106 31.9 24.9 37.3

Manual Integrations
APPROVED



Reviewed By :Krupa Patel 12/06/2022
Supervised By :Mahesh Dadoda 12/06/2022




#68
m/p-Xylenes
Concen: 41.699 ug/l
RT: 12.070 min Scan# 1703
Delta R.T. -0.001 min
Lab File: VN075558.D
Acq: 05 Dec 2022 14:55



Tgt Ion:106 Resp: 147054
Ion Ratio Lower Upper
106 100
91 193.9 156.0 234.0









#69
o-Xylene
Concen: 21.149 ug/l
RT: 12.400 min Scan# 1759
Delta R.T. -0.000 min
Lab File: VN075558.D
Acq: 05 Dec 2022 14:55

Instrument :
MSVOA_N
ClientSampleId :
VN1205WBSD01

Tgt Ion:106 Resp: 7410
Ion Ratio Lower Upper
106 100
91 203.0 105.2 315.5




Manual Integrations
APPROVED


Reviewed By :Krupa Patel 12/06/2022
Supervised By :Mahesh Dadoda 12/06/2022



#70
Styrene
Concen: 21.711 ug/l
RT: 12.418 min Scan# 1762
Delta R.T. 0.005 min
Lab File: VN075558.D
Acq: 05 Dec 2022 14:55


Tgt Ion:104 Resp: 120009
Ion Ratio Lower Upper
104 100
78 45.9 38.1 57.1
103 52.8 44.6 66.8





#71
Bromoform
Concen: 23.158 ug/l
RT: 12.582 min Scan# 1789
Delta R.T. 0.006 min
Lab File: VN075558.D
Acq: 05 Dec 2022 14:55




Instrument :
MSVOA_N
ClientSampleId :
VN1205WBSD01



Tgt Ion:173 Resp: 33484
Ion Ratio Lower Upper
173 100
175 47.9 24.6 74.0
254 0.0 0.0 0.0




Manual Integrations
APPROVED


Reviewed By :Krupa Patel 12/06/2022
Supervised By :Mahesh Dadoda 12/06/2022



#72
1,4-Dichlorobenzene-d4
Concen: 50.000 ug/l
RT: 13.794 min Scan# 1996
Delta R.T. -0.000 min
Lab File: VN075558.D
Acq: 05 Dec 2022 14:55

Tgt Ion:152 Resp: 126617
Ion Ratio Lower Upper
152 100
115 56.7 30.2 90.6
150 161.6 0.0 348.4









#73
Isopropylbenzene
Concen: 20.328 ug/l
RT: 12.700 min Scan# 1809
Delta R.T. 0.006 min
Lab File: VN075558.D
Acq: 05 Dec 2022 14:55

Instrument :
MSVOA_N
ClientSampleId :
VN1205WBSD01

Tgt Ion:105 Resp: 18804
Ion Ratio Lower Upper
105 100
120 26.6 13.1 39.1




Manual Integrations
APPROVED


Reviewed By :Krupa Patel 12/06/2022
Supervised By :Mahesh Dadoda 12/06/2022



#74
N-ethyl acetate
Concen: 19.527 ug/l
RT: 12.500 min Scan# 1776
Delta R.T. 0.006 min
Lab File: VN075558.D
Acq: 05 Dec 2022 14:55

Tgt Ion: 43 Resp: 55535
Ion Ratio Lower Upper
43 100
70 50.4 39.8 59.8
55 26.5 19.9 29.9
61 25.7 21.7 32.5









#75
1,1,2,2-Tetrachloroethane
Concen: 20.514 ug/l
RT: 12.941 min Scan# 1851
Delta R.T. 0.000 min
Lab File: VN075558.D
Acq: 05 Dec 2022 14:55

Instrument :
MSVOA_N
ClientSampleId :
VN1205WBSD01

Tgt Ion: 83 Resp: 57850
Ion Ratio Lower Upper
83 100
131 10.1 5.1 15.3
85 63.5 31.6 94.7




Manual Integrations
APPROVED


Reviewed By :Krupa Patel 12/06/2022
Supervised By :Mahesh Dadoda 12/06/2022



#76
1,2,3-Trichloropropane
Concen: 20.142 ug/l m
RT: 12.994 min Scan# 1860
Delta R.T. -0.000 min
Lab File: VN075558.D
Acq: 05 Dec 2022 14:55


Tgt Ion: 75 Resp: 48220
Ion Ratio Lower Upper
75 100
77 178.3 75.6 226.8





#77
Bromobenzene
Concen: 20.833 ug/l
RT: 12.982 min Scan# 1858
Delta R.T. 0.000 min
Lab File: VN075558.D
Acq: 05 Dec 2022 14:55




Instrument :
MSVOA_N
ClientSampleId :
VN1205WBSD01




Tgt Ion:156 Resp: 47578
Ion Ratio Lower Upper
156 100
77 180.8 108.5 325.5
158 94.7 47.7 143.1

Manual Integrations
APPROVED



Reviewed By :Krupa Patel 12/06/2022
Supervised By :Mahesh Dadoda 12/06/2022




#78
n-propylbenzene
Concen: 20.562 ug/l
RT: 13.041 min Scan# 1868
Delta R.T. 0.006 min
Lab File: VN075558.D
Acq: 05 Dec 2022 14:55



Tgt Ion: 91 Resp: 212278
Ion Ratio Lower Upper
91 100
120 24.2 11.9 35.5





#79

2-Chlorotoluene

Concen: 20.624 ug/l

RT: 13.129 min Scan# 1882

Delta R.T. 0.005 min

Lab File: VN075558.D

Acq: 05 Dec 2022 14:55

Instrument :

MSVOA_N

ClientSampleId :

VN1205WBSD01

Tgt Ion: 91 Resp: 12828

Ion Ratio Lower Upper

91 100





126 35.6 16.9 50.7

Manual Integrations

APPROVED

Reviewed By :Krupa Patel 12/06/2022

Supervised By :Mahesh Dadoda 12/06/2022



#80

1,3,5-Trimethylbenzene

Concen: 20.977 ug/l

RT: 13.176 min Scan# 1891

Delta R.T. 0.000 min

Lab File: VN075558.D




Acq: 05 Dec 2022 14:55


Tgt Ion:105 Resp: 161221

Ion Ratio Lower Upper

105 100


120 50.4 24.7 74.1





#81
trans-1,4-Dichloro-2-butene
Concen: 22.940 ug/l
RT: 12.741 min Scan# 1816
Delta R.T. 0.006 min
Lab File: VN075558.D
Acq: 05 Dec 2022 14:55




Instrument :
MSVOA_N
ClientSampleId :
VN1205WBSD01



Tgt Ion: 75 Resp: 1856
Ion Ratio Lower Upper
75 100
53 82.5 78.5 117.7
89 40.5 37.7 56.5




Manual Integrations
APPROVED


Reviewed By :Krupa Patel 12/06/2022
Supervised By :Mahesh Dadoda 12/06/2022



#82
4-Chlorotoluene
Concen: 20.624 ug/l
RT: 13.129 min Scan# 1883
Delta R.T. 0.005 min
Lab File: VN075558.D
Acq: 05 Dec 2022 14:55


Tgt Ion: 91 Resp: 128287
Ion Ratio Lower Upper
91 100
126 35.6 16.9 50.7





#83
tert-Butylbenzene
Concen: 20.435 ug/l
RT: 13.441 min Scan# 1936
Delta R.T. -0.000 min
Lab File: VN075558.D
Acq: 05 Dec 2022 14:55




Instrument :
MSVOA_N
ClientSampleId :
VN1205WBSD01




Tgt Ion:119 Resp: 143190
Ion Ratio Lower Upper
119 100
91 63.0 32.8 98.4
134 28.0 12.3 36.9

Manual Integrations
APPROVED



Reviewed By :Krupa Patel 12/06/2022
Supervised By :Mahesh Dadoda 12/06/2022




#84
1,2,4-Trimethylbenzene
Concen: 21.437 ug/l
RT: 13.488 min Scan# 1944
Delta R.T. 0.006 min
Lab File: VN075558.D
Acq: 05 Dec 2022 14:55



Tgt Ion:105 Resp: 165255
Ion Ratio Lower Upper
105 100
120 46.1 22.7 68.1









#85
 sec-Butylbenzene
 Concen: 20.839 ug/l
 RT: 13.617 min Scan# 1966
 Delta R.T. -0.001 min
 Lab File: VN075558.D
 Acq: 05 Dec 2022 14:55

Instrument :
 MSVOA_N
 ClientSampleId :
 VN1205WBSD01

Tgt Ion:105 Resp: 198139
 Ion Ratio Lower Upper
 105 100
 134 20.8 10.2 30.6




Manual Integrations
 APPROVED

Reviewed By :Krupa Patel 12/06/2022
 Supervised By :Mahesh Dadoda 12/06/2022



#86
 p-Isopropyltoluene
 Concen: 21.121 ug/l
 RT: 13.735 min Scan# 1986
 Delta R.T. 0.006 min
 Lab File: VN075558.D
 Acq: 05 Dec 2022 14:55


Tgt Ion:119 Resp: 168958
 Ion Ratio Lower Upper
 119 100
 134 26.3 13.1 39.3
 91 22.8 11.9 35.5





#87
1,3-Dichlorobenzene
Concen: 20.209 ug/l
RT: 13.735 min Scan# 1986
Delta R.T. 0.000 min
Lab File: VN075558.D
Acq: 05 Dec 2022 14:55




Instrument :
MSVOA_N
ClientSampleId :
VN1205WBSD01




Tgt Ion:146 Resp: 87449
Ion Ratio Lower Upper
146 100
111 39.7 20.8 62.5
148 64.9 31.1 93.5

Manual Integrations
APPROVED



Reviewed By :Krupa Patel 12/06/2022
Supervised By :Mahesh Dadoda 12/06/2022




#88
1,4-Dichlorobenzene
Concen: 19.891 ug/l
RT: 13.817 min Scan# 2000
Delta R.T. 0.005 min
Lab File: VN075558.D
Acq: 05 Dec 2022 14:55



Tgt Ion:146 Resp: 87432
Ion Ratio Lower Upper
146 100
111 39.0 20.1 60.3
148 63.4 31.9 95.7









#89
n-Butylbenzene
Concen: 19.620 ug/l
RT: 14.059 min Scan# 2041
Delta R.T. -0.000 min
Lab File: VN075558.D
Acq: 05 Dec 2022 14:55

Instrument :
MSVOA_N
ClientSampleId :
VN1205WBSD01

Tgt Ion: 91 Resp: 132558
Ion Ratio Lower Upper
91 100
92 53.4 26.5 79.4
134 27.6 13.0 39.0




Manual Integrations
APPROVED


Reviewed By :Krupa Patel 12/06/2022
Supervised By :Mahesh Dadoda 12/06/2022



#90
Hexachloroethane
Concen: 21.775 ug/l
RT: 14.335 min Scan# 2088
Delta R.T. 0.000 min
Lab File: VN075558.D
Acq: 05 Dec 2022 14:55

Tgt Ion:117 Resp: 30050
Ion Ratio Lower Upper
117 100
201 77.8 31.5 94.5









#91
1,2-Dichlorobenzene
Concen: 20.979 ug/l
RT: 14.106 min Scan# 2049
Delta R.T. -0.000 min
Lab File: VN075558.D
Acq: 05 Dec 2022 14:55

Instrument :
MSVOA_N
ClientSampleId :
VN1205WBSD01

Tgt Ion:146 Resp: 8848
Ion Ratio Lower Upper
146 100
111 40.6 21.3 63.9
148 64.9 31.4 94.2




Manual Integrations
APPROVED

Reviewed By :Krupa Patel 12/06/2022
Supervised By :Mahesh Dadoda 12/06/2022



#92
1,2-Dibromo-3-Chloropropane
Concen: 20.405 ug/l
RT: 14.723 min Scan# 2154
Delta R.T. 0.000 min
Lab File: VN075558.D
Acq: 05 Dec 2022 14:55

Tgt Ion: 75 Resp: 10901
Ion Ratio Lower Upper
75 100
155 97.1 42.1 126.3
157 121.0 51.7 155.1





#93

1,2,4-Trichlorobenzene

Concen: 20.830 ug/l

RT: 15.400 min Scan# 21

Delta R.T. 0.006 min

Lab File: VN075558.D

Acq: 05 Dec 2022 14:55

Instrument :

MSVOA_N

ClientSampleId :

VN1205WBSD01





Tgt Ion:180	Resp:	45024
Ion Ratio	Lower	Upper
180	100	
182	93.6	47.8 143.4
145	31.5	16.8 50.3

Manual Integrations

APPROVED

Reviewed By :Krupa Patel 12/06/2022

Supervised By :Mahesh Dadoda 12/06/2022



#94

Hexachlorobutadiene

Concen: 21.225 ug/l




RT: 15.500 min Scan# 2286


Delta R.T. -0.006 min

Lab File: VN075558.D

Acq: 05 Dec 2022 14:55


Tgt Ion:225	Resp:	23896
Ion Ratio	Lower	Upper
225	100	
223	60.7	32.3 96.9
227	63.4	32.0 96.2





#95
Naphthalene
Concen: 19.254 ug/l
RT: 15.647 min Scan# 2311
Delta R.T. 0.006 min
Lab File: VN075558.D
Acq: 05 Dec 2022 14:55




Instrument :
MSVOA_N
ClientSampleId :
VN1205WBSD01




Tgt Ion:128 Resp: 13092
Ion Ratio Lower Upper
128 100
127 13.0 10.4 15.6
129 10.4 8.6 12.8

Manual Integrations
APPROVED

Reviewed By :Krupa Patel 12/06/2022
Supervised By :Mahesh Dadoda 12/06/2022



#96
1,2,3-Trichlorobenzene
Concen: 20.573 ug/l
RT: 15.847 min Scan# 2345
Delta R.T. 0.006 min
Lab File: VN075558.D
Acq: 05 Dec 2022 14:55



Tgt Ion:180 Resp: 43705
Ion Ratio Lower Upper
180 100
182 94.9 47.3 141.8
145 31.9 17.6 52.9

