

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN120722\
 Data File : VN075614.D
 Acq On : 07 Dec 2022 18:19
 Operator : JC\MD
 Sample : PB149441TB
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB149441TB

Quant Time: Dec 08 00:22:58 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N112122W.M
 Quant Title : SW846 8260
 QLast Update : Wed Nov 23 00:18:03 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.230	168	199588	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	331219	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.871	117	304479	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	138901	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.583	65	55529	51.830	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	103.660%
35) Dibromofluoromethane	8.171	113	69790	53.638	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	107.280%
50) Toluene-d8	10.571	98	152434	52.016	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	104.040%
62) 4-Bromofluorobenzene	12.853	95	112263	51.900	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	103.800%
Target Compounds						
					Qvalue	
16) Acetone	4.442	43	20828	20.842	ug/l #	87
20) Methylene Chloride	5.283	84	10242	3.874	ug/l #	89
37) Ethyl Acetate	7.565	43	26889	10.524	ug/l	98
43) Isopropyl Acetate	8.694	43	128477	31.328	ug/l #	92

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN120722\
Data File : VN075614.D
Acq On : 07 Dec 2022 18:19
Operator : JC\MD
Sample : PB149441TB
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 18 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB149441TB

Quant Time: Dec 08 00:22:58 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N112122W.M
Quant Title : SW846 8260
QLast Update : Wed Nov 23 00:18:03 2022
Response via : Initial Calibration

