

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN120722\
 Data File : VN075615.D
 Acq On : 07 Dec 2022 18:43
 Operator : JC\MD
 Sample : N5883-02
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 MH-18

Quant Time: Dec 08 00:23:10 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N112122W.M
 Quant Title : SW846 8260
 QLast Update : Wed Nov 23 00:18:03 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.230	168	199522	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	334540	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.870	117	300073	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	136975	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.588	65	56991	53.213	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	= 106.420%	
35) Dibromofluoromethane	8.177	113	72380	55.076	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 110.160%	
50) Toluene-d8	10.571	98	159631	53.931	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	= 107.860%	
62) 4-Bromofluorobenzene	12.853	95	115154	52.709	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	= 105.420%	
Target Compounds						
					Qvalue	
16) Acetone	4.447	43	22760	22.783	ug/l #	83
20) Methylene Chloride	5.289	84	10531	4.023	ug/l	86
37) Ethyl Acetate	7.565	43	25435	9.856	ug/l	98
43) Isopropyl Acetate	8.694	43	170068	41.058	ug/l #	82

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN120722\
Data File : VN075615.D
Acq On : 07 Dec 2022 18:43
Operator : JC\MD
Sample : N5883-02
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 19 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
MH-18

Quant Time: Dec 08 00:23:10 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N112122W.M
Quant Title : SW846 8260
QLast Update : Wed Nov 23 00:18:03 2022
Response via : Initial Calibration

