

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN120722\
 Data File : VN075616.D
 Acq On : 07 Dec 2022 19:07
 Operator : JC\MD
 Sample : N5895-02
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 MH-19

Quant Time: Dec 08 00:23:21 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N112122W.M
 Quant Title : SW846 8260
 QLast Update : Wed Nov 23 00:18:03 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.230	168	199783	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	341557	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.871	117	306909	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	140477	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.588	65	56804	52.969	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	= 105.940%	
35) Dibromofluoromethane	8.171	113	72096	53.733	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 107.460%	
50) Toluene-d8	10.571	98	154811	51.228	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	= 102.460%	
62) 4-Bromofluorobenzene	12.853	95	115681	51.862	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	= 103.720%	
Target Compounds						
					Qvalue	
16) Acetone	4.453	43	22008	22.002	ug/l	94
20) Methylene Chloride	5.289	84	11556	4.534	ug/l	87
37) Ethyl Acetate	7.571	43	25722	9.763	ug/l	99
43) Isopropyl Acetate	8.694	43	236546	55.934	ug/l #	75

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN120722\
Data File : VN075616.D
Acq On : 07 Dec 2022 19:07
Operator : JC\MD
Sample : N5895-02
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 20 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
MH-19

Quant Time: Dec 08 00:23:21 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N112122W.M
Quant Title : SW846 8260
QLast Update : Wed Nov 23 00:18:03 2022
Response via : Initial Calibration

