

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN120722\
 Data File : VN075617.D
 Acq On : 07 Dec 2022 19:31
 Operator : JC\MD
 Sample : N5897-18
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 21G4

Quant Time: Dec 08 00:23:32 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N112122W.M
 Quant Title : SW846 8260
 QLast Update : Wed Nov 23 00:18:03 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.230	168	195488	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	332353	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.870	117	303303	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	134576	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.582	65	57191	54.501	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	= 109.000%	
35) Dibromofluoromethane	8.177	113	71815	55.006	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 110.020%	
50) Toluene-d8	10.571	98	158294	53.831	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	= 107.660%	
62) 4-Bromofluorobenzene	12.853	95	114558	52.781	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	= 105.560%	
Target Compounds						
					Qvalue	
16) Acetone	4.447	43	33401	34.125	ug/l	99
20) Methylene Chloride	5.289	84	11208	4.483	ug/l	91
25) 2-Butanone	7.500	43	7000	5.123	ug/l	98
37) Ethyl Acetate	7.577	43	24751	9.654	ug/l	97
43) Isopropyl Acetate	8.694	43	182746	44.409	ug/l #	79

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN120722\
Data File : VN075617.D
Acq On : 07 Dec 2022 19:31
Operator : JC\MD
Sample : N5897-18
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 21 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
21G4

Quant Time: Dec 08 00:23:32 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N112122W.M
Quant Title : SW846 8260
QLast Update : Wed Nov 23 00:18:03 2022
Response via : Initial Calibration

