

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN120722\
 Data File : VN075619.D
 Acq On : 07 Dec 2022 20:19
 Operator : JC\MD
 Sample : N5897-20
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 23 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 2997

Quant Time: Dec 08 00:23:56 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N112122W.M
 Quant Title : SW846 8260
 QLast Update : Wed Nov 23 00:18:03 2022
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.229	168	193503	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	325923	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.870	117	294211	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	137020	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.582	65	54595	52.561	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	105.120%
35) Dibromofluoromethane	8.177	113	69481	54.268	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	108.540%
50) Toluene-d8	10.571	98	150569	52.214	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	104.420%
62) 4-Bromofluorobenzene	12.853	95	111433	52.354	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	104.700%
Target Compounds						
						Qvalue
16) Acetone	4.447	43	33830	34.918	ug/l	99
20) Methylene Chloride	5.283	84	11143	4.508	ug/l #	79
25) 2-Butanone	7.500	43	7492	5.539	ug/l	91
37) Ethyl Acetate	7.571	43	24382	9.698	ug/l #	93
43) Isopropyl Acetate	8.694	43	185035	45.852	ug/l #	79

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN120722\
Data File : VN075619.D
Acq On : 07 Dec 2022 20:19
Operator : JC\MD
Sample : N5897-20
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 23 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
2997

Quant Time: Dec 08 00:23:56 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N112122W.M
Quant Title : SW846 8260
QLast Update : Wed Nov 23 00:18:03 2022
Response via : Initial Calibration

