

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN120722\
 Data File : VN075620.D
 Acq On : 07 Dec 2022 20:43
 Operator : JC\MD
 Sample : N5897-21
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 24 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 STOCK-PILE

Quant Time: Dec 08 00:24:08 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N112122W.M
 Quant Title : SW846 8260
 QLast Update : Wed Nov 23 00:18:03 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.230	168	182594	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	313491	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.871	117	280631	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	123672	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.588	65	55902	57.035	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	114.060%
35) Dibromofluoromethane	8.177	113	70002	56.843	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	113.680%
50) Toluene-d8	10.571	98	144337	52.038	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	104.080%
62) 4-Bromofluorobenzene	12.853	95	104333	50.962	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	101.920%
Target Compounds						
					Qvalue	
16) Acetone	4.447	43	29415	32.175	ug/l	97
20) Methylene Chloride	5.289	84	11069	4.815	ug/l	87
37) Ethyl Acetate	7.571	43	24316	10.055	ug/l	100
43) Isopropyl Acetate	8.694	43	168308	43.361	ug/l #	80

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN120722\
Data File : VN075620.D
Acq On : 07 Dec 2022 20:43
Operator : JC\MD
Sample : N5897-21
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 24 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
STOCK-PILE

Quant Time: Dec 08 00:24:08 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N112122W.M
Quant Title : SW846 8260
QLast Update : Wed Nov 23 00:18:03 2022
Response via : Initial Calibration

