

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN120722\
 Data File : VN075626.D
 Acq On : 07 Dec 2022 23:08
 Operator : JC\MD
 Sample : VN1207WBL02
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 30 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VN1207WBL02

Quant Time: Dec 08 00:25:30 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N112122W.M
 Quant Title : SW846 8260
 QLast Update : Wed Nov 23 00:18:03 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.229	168	179617	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	298976	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.870	117	257796	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	116366	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.582	65	48160	49.950	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	99.900%
35) Dibromofluoromethane	8.177	113	62718	53.401	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	106.800%
50) Toluene-d8	10.571	98	127165	48.073	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	96.140%
62) 4-Bromofluorobenzene	12.853	95	93857	48.071	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	96.140%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN120722\
Data File : VN075626.D
Acq On : 07 Dec 2022 23:08
Operator : JC\MD
Sample : VN1207WBL02
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 30 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VN1207WBL02

Quant Time: Dec 08 00:25:30 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N112122W.M
Quant Title : SW846 8260
QLast Update : Wed Nov 23 00:18:03 2022
Response via : Initial Calibration

