

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN120722\
 Data File : VN075630.D
 Acq On : 08 Dec 2022 00:44
 Operator : JC\MD
 Sample : PB149441ZHE02
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 34 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB149441ZHE02

Quant Time: Dec 08 01:20:34 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N112122W.M
 Quant Title : SW846 8260
 QLast Update : Wed Nov 23 00:18:03 2022
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.230	168	188537	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	323188	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.870	117	292488	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	131157	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.582	65	53598	52.960	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	= 105.920%	
35) Dibromofluoromethane	8.171	113	67815	53.415	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 106.820%	
50) Toluene-d8	10.571	98	149169	52.167	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	= 104.340%	
62) 4-Bromofluorobenzene	12.853	95	106815	50.609	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	= 101.220%	
Target Compounds						
					Qvalue	
16) Acetone	4.442	43	15360	16.271	ug/l	99
20) Methylene Chloride	5.289	84	6961	2.420	ug/l	95
37) Ethyl Acetate	7.571	43	24683	9.901	ug/l	97
43) Isopropyl Acetate	8.694	43	288318	72.051	ug/l #	72

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN120722\
Data File : VN075630.D
Acq On : 08 Dec 2022 00:44
Operator : JC\MD
Sample : PB149441ZHE02
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 34 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB149441ZHE02

Quant Time: Dec 08 01:20:34 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N112122W.M
Quant Title : SW846 8260
QLast Update : Wed Nov 23 00:18:03 2022
Response via : Initial Calibration

