

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN120722\
 Data File : VN075634.D
 Acq On : 08 Dec 2022 02:20
 Operator : JC\MD
 Sample : PB149441ZHE06
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 38 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB149441ZHE06

Quant Time: Dec 08 06:45:09 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N112122W.M
 Quant Title : SW846 8260
 QLast Update : Wed Nov 23 00:18:03 2022
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.230	168	188052	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	324433	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.871	117	294932	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	133484	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.583	65	53598	53.097	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	= 106.200%	
35) Dibromofluoromethane	8.171	113	67222	52.745	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 105.480%	
50) Toluene-d8	10.571	98	148784	51.832	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	= 103.660%	
62) 4-Bromofluorobenzene	12.853	95	107858	50.907	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	= 101.820%	
Target Compounds						
					Qvalue	
16) Acetone	4.453	43	15743	16.720	ug/l	97
18) Methyl Acetate	5.042	43	6367	1.424	ug/l	94
20) Methylene Chloride	5.289	84	7336	2.631	ug/l #	86
37) Ethyl Acetate	7.571	43	25723	10.279	ug/l	98
43) Isopropyl Acetate	8.694	43	266113	66.247	ug/l #	74

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN120722\
Data File : VN075634.D
Acq On : 08 Dec 2022 02:20
Operator : JC\MD
Sample : PB149441ZHE06
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 38 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB149441ZHE06

Quant Time: Dec 08 06:45:09 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N112122W.M
Quant Title : SW846 8260
QLast Update : Wed Nov 23 00:18:03 2022
Response via : Initial Calibration

