

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN120722\
 Data File : VN075635.D
 Acq On : 08 Dec 2022 02:45
 Operator : JC\MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 39 Sample Multiplier: 1

Instrument :
 MSVOA_N
 LabSampleId :
 VSTDCCC050

Quant Time: Dec 08 06:45:18 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N112122W.M
 Quant Title : SW846 8260
 QLast Update : Wed Nov 23 00:18:03 2022
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	1.000	1.000	0.0	89	0.00
2 T	Dichlorodifluoromethane	0.382	0.344	9.9	81	0.00
3 P	Chloromethane	0.476	0.533	-12.0	108	0.00
4 C	Vinyl Chloride	0.600	0.614	-2.3#	94	0.00
5 T	Bromomethane	0.527	0.655	-24.3	105	0.00
6 T	Chloroethane	0.454	0.510	-12.3	107	0.00
7 T	Trichlorofluoromethane	0.857	0.750	12.5	80	-0.01
8 T	Diethyl Ether	0.300	0.435	-45.0#	124	0.00
9 T	1,1,2-Trichlorotrifluoroeth	0.473	0.431	8.9	82	0.00
10 T	Methyl Iodide	0.766	0.979	-27.8#	114	0.00
11 T	Tert butyl alcohol	0.113	0.146	-29.2#	126	0.00
12 CM	1,1-Dichloroethene	0.463	0.469	-1.3#	91	0.00
13 T	Acrolein	0.098	0.135	-37.8#	128	0.00
14 T	Allyl chloride	0.599	0.710	-18.5	114	0.00
15 T	Acrylonitrile	0.258	0.331	-28.3#	116	0.00
16 T	Acetone	0.250	0.260	-4.0	100	0.00
17 T	Carbon Disulfide	1.203	1.178	2.1	88	0.00
18 T	Methyl Acetate	0.741	0.888	-19.8	123	0.00
19 T	Methyl tert-butyl Ether	1.640	2.240	-36.6#	125	0.00
20 T	Methylene Chloride	0.587	0.693	-18.1	119	0.00
21 T	trans-1,2-Dichloroethene	0.528	0.574	-8.7	100	0.00
22 T	Diisopropyl ether	1.367	1.821	-33.2#	118	0.01
23 T	Vinyl Acetate	1.087	1.502	-38.2#	119	0.00
24 P	1,1-Dichloroethane	0.877	1.070	-22.0	108	0.00
25 T	2-Butanone	0.350	0.416	-18.9	111	0.00
26 T	2,2-Dichloropropane	0.815	0.624	23.4	69	0.00
27 T	cis-1,2-Dichloroethene	0.597	0.758	-27.0#	110	0.00
28 T	Bromochloromethane	0.356	0.485	-36.2#	124	0.00
29 T	Tetrahydrofuran	0.213	0.274	-28.6#	115	0.00
30 C	Chloroform	1.007	1.228	-21.9#	113	0.00
31 T	Cyclohexane	0.853	0.664	22.2	74	0.00
32 T	1,1,1-Trichloroethane	0.915	0.955	-4.4	93	0.00
33 S	1,2-Dichloroethane-d4	0.268	0.343	-28.0#	115	0.00
34 I	1,4-Difluorobenzene	1.000	1.000	0.0	96	0.00
35 S	Dibromofluoromethane	0.196	0.245	-25.0	114	0.00
36 T	1,1-Dichloropropene	0.428	0.389	9.1	86	0.00
37 T	Ethyl Acetate	0.386	0.472	-22.3	112	0.00
38 T	Carbon Tetrachloride	0.471	0.453	3.8	89	0.00
39 T	Methylcyclohexane	0.536	0.449	16.2	78	0.00
40 TM	Benzene	1.311	1.423	-8.5	104	0.00
41 T	Methacrylonitrile	0.197	0.246	-24.9	121	0.00
42 TM	1,2-Dichloroethane	0.439	0.539	-22.8	119	0.00
43 T	Isopropyl Acetate	0.619	0.992	-60.3#	149	0.00
44 TM	Trichloroethene	0.359	0.359	0.0	97	0.00
45 C	1,2-Dichloropropane	0.313	0.368	-17.6#	113	0.00
46 T	Dibromomethane	0.236	0.293	-24.2	119	0.00
47 T	Bromodichloromethane	0.458	0.564	-23.1	116	0.00
48 T	Methyl methacrylate	0.286	0.363	-26.9#	119	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN120722\
 Data File : VN075635.D
 Acq On : 08 Dec 2022 02:45
 Operator : JC\MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 39 Sample Multiplier: 1

Instrument :
 MSVOA_N
 LabSampleId :
 VSTDCCC050

Quant Time: Dec 08 06:45:18 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N112122W.M
 Quant Title : SW846 8260
 QLast Update : Wed Nov 23 00:18:03 2022
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
49 T	1,4-Dioxane	0.008	0.010	-25.0	111	0.00
50 S	Toluene-d8	0.442	0.454	-2.7	93	0.00
51 T	4-Methyl-2-Pentanone	0.394	0.501	-27.2#	120	0.00
52 CM	Toluene	0.845	0.898	-6.3#	98	0.00
53 T	t-1,3-Dichloropropene	0.460	0.572	-24.3	109	0.00
54 T	cis-1,3-Dichloropropene	0.499	0.591	-18.4	107	0.00
55 T	1,1,2-Trichloroethane	0.344	0.427	-24.1	121	0.00
56 T	Ethyl methacrylate	0.465	0.633	-36.1#	120	0.00
57 T	1,3-Dichloropropane	0.561	0.683	-21.7	117	0.00
58 T	2-Chloroethyl Vinyl ether	0.168	0.199	-18.5	102	0.00
59 T	2-Hexanone	0.296	0.375	-26.7#	117	0.00
60 T	Dibromochloromethane	0.353	0.481	-36.3#	117	0.00
61 T	1,2-Dibromoethane	0.357	0.448	-25.5#	117	0.00
62 S	4-Bromofluorobenzene	0.327	0.357	-9.2	97	0.00
63 I	Chlorobenzene-d5	1.000	1.000	0.0	94	0.00
64 T	Tetrachloroethene	0.382	0.316	17.3	82	0.00
65 PM	Chlorobenzene	1.033	1.107	-7.2	101	0.00
66 T	1,1,1,2-Tetrachloroethane	0.385	0.454	-17.9	109	0.00
67 C	Ethyl Benzene	1.762	1.782	-1.1#	91	0.00
68 T	m/p-Xylenes	0.715	0.730	-2.1	93	0.00
69 T	o-Xylene	0.711	0.751	-5.6	99	0.00
70 T	Styrene	1.121	1.268	-13.1	100	0.00
71 P	Bromoform	0.293	0.404	-37.9#	118	0.00
72 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	92	0.00
73 T	Isopropylbenzene	3.653	3.452	5.5	89	0.00
74 T	N-amyl acetate	1.123	1.531	-36.3#	122	0.00
75 P	1,1,2,2-Tetrachloroethane	1.114	1.363	-22.4	119	0.00
76 T	1,2,3-Trichloropropane	0.945	1.050	-11.1	109	0.00
77 T	Bromobenzene	0.902	0.997	-10.5	106	0.00
78 T	n-propylbenzene	4.077	3.837	5.9	86	0.00
79 T	2-Chlorotoluene	2.456	2.446	0.4	94	0.00
80 T	1,3,5-Trimethylbenzene	3.035	3.052	-0.6	91	0.00
81 T	trans-1,4-Dichloro-2-butene	0.320	0.364	-13.7	100	0.00
82 T	4-Chlorotoluene	2.456	2.446	0.4	94	0.00
83 T	tert-Butylbenzene	2.767	2.585	6.6	87	0.00
84 T	1,2,4-Trimethylbenzene	3.044	3.107	-2.1	92	0.00
85 T	sec-Butylbenzene	3.755	3.541	5.7	84	0.00
86 T	p-Isopropyltoluene	3.159	3.111	1.5	86	0.00
87 T	1,3-Dichlorobenzene	1.709	1.744	-2.0	95	0.00
88 T	1,4-Dichlorobenzene	1.736	1.744	-0.5	96	0.00
89 T	n-Butylbenzene	2.668	2.431	8.9	81	0.00
90 T	Hexachloroethane	0.545	0.535	1.8	85	0.00
91 T	1,2-Dichlorobenzene	1.666	1.825	-9.5	99	0.00
92 T	1,2-Dibromo-3-Chloropropane	0.211	0.265	-25.6#	110	0.00
93 T	1,2,4-Trichlorobenzene	0.854	0.911	-6.7	90	0.00
94 T	Hexachlorobutadiene	0.445	0.435	2.2	85	0.00
95 T	Naphthalene	2.685	3.125	-16.4	99	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN120722\
Data File : VN075635.D
Acq On : 08 Dec 2022 02:45
Operator : JC\MD
Sample : VSTDCCC050
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 39 Sample Multiplier: 1

Instrument :
MSVOA_N
LabSampleId :
VSTDCCC050

Quant Time: Dec 08 06:45:18 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N112122W.M
Quant Title : SW846 8260
QLast Update : Wed Nov 23 00:18:03 2022
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
96 T	1,2,3-Trichlorobenzene	0.839	0.911	-8.6	95	0.00

(#) = Out of Range SPCC's out = 0 CCC's out = 6