

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN100722\
 Data File : VN074805.D
 Acq On : 07 Oct 2022 12:42
 Operator : JC\MD
 Sample : N4974-01
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 257248-6

Quant Time: Oct 08 03:51:31 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N100522W.M
 Quant Title : SW846 8260
 QLast Update : Thu Oct 06 05:37:50 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.235	168	251835	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	9.106	114	472108	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.871	117	462875	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	235753	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.588	65	405897	68.023	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	= 136.040%#	
35) Dibromofluoromethane	8.177	113	319098	66.897	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 133.800%#	
50) Toluene-d8	10.571	98	1292571	63.153	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	= 126.300%#	
62) 4-Bromofluorobenzene	12.853	95	325251	52.276	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	= 104.560%	

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN100722\
Data File : VN074805.D
Acq On : 07 Oct 2022 12:42
Operator : JC\MD
Sample : N4974-01
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
257248-6

Quant Time: Oct 08 03:51:31 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N100522W.M
Quant Title : SW846 8260
QLast Update : Thu Oct 06 05:37:50 2022
Response via : Initial Calibration

