

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN100722\
 Data File : VN074807.D
 Acq On : 07 Oct 2022 13:30
 Operator : JC\MD
 Sample : N4979-01
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 251032-4

Quant Time: Oct 07 13:48:10 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N100522W.M
 Quant Title : SW846 8260
 QLast Update : Thu Oct 06 05:37:50 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.235	168	259145	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.112	114	476940	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	475532	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	230362	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.588	65	396555	64.583	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	129.160%#
35) Dibromofluoromethane	8.182	113	301229	62.511	ug/l	0.01
Spiked Amount	50.000	Range	75 - 124	Recovery	=	125.020%#
50) Toluene-d8	10.571	98	1227402	59.361	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	118.720%#
62) 4-Bromofluorobenzene	12.853	95	304825	48.497	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	97.000%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN100722\
Data File : VN074807.D
Acq On : 07 Oct 2022 13:30
Operator : JC\MD
Sample : N4979-01
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
251032-4

Quant Time: Oct 07 13:48:10 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N100522W.M
Quant Title : SW846 8260
QLast Update : Thu Oct 06 05:37:50 2022
Response via : Initial Calibration

