

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN100722\
 Data File : VN074818.D
 Acq On : 07 Oct 2022 17:56
 Operator : JC\MD
 Sample : N5009-02
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 22 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 006-4

Quant Time: Oct 08 03:54:59 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N100522W.M
 Quant Title : SW846 8260
 QLast Update : Thu Oct 06 05:37:50 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.236	168	212538	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	411245	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.871	117	415437	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	199854	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.588	65	330397	65.608	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	131.220%#
35) Dibromofluoromethane	8.177	113	256836	61.813	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	123.620%
50) Toluene-d8	10.571	98	992950	55.694	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	111.380%
62) 4-Bromofluorobenzene	12.853	95	259480	47.877	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	95.760%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN100722\
Data File : VN074818.D
Acq On : 07 Oct 2022 17:56
Operator : JC\MD
Sample : N5009-02
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 22 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
006-4

Quant Time: Oct 08 03:54:59 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N100522W.M
Quant Title : SW846 8260
QLast Update : Thu Oct 06 05:37:50 2022
Response via : Initial Calibration

