

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN100722\
 Data File : VN074805.D
 Acq On : 07 Oct 2022 12:42
 Operator : JC\MD
 Sample : N4974-01
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 257248-6

Quant Time: Oct 08 03:51:31 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N100522W.M
 Quant Title : SW846 8260
 QLast Update : Thu Oct 06 05:37:50 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.235	168	251835	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	9.106	114	472108	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.871	117	462875	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	235753	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.588	65	405897	68.023	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	136.040%#
35) Dibromofluoromethane	8.177	113	319098	66.897	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	133.800%#
50) Toluene-d8	10.571	98	1292571	63.153	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	126.300%#
62) 4-Bromofluorobenzene	12.853	95	325251	52.276	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	104.560%


Target Compounds	Qvalue


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN100722\
Data File : VN074805.D
Acq On : 07 Oct 2022 12:42
Operator : JC\MD
Sample : N4974-01
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
257248-6

Quant Time: Oct 08 03:51:31 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N100522W.M
Quant Title : SW846 8260
QLast Update : Thu Oct 06 05:37:50 2022
Response via : Initial Calibration









#1
Pentafluorobenzene
Concen: 50.000 ug/l
RT: 8.235 min Scan# 1051
Delta R.T. -0.001 min
Lab File: VN074805.D
Acq: 07 Oct 2022 12:42




Instrument :
MSVOA_N
ClientSampleId :
257248-6


Tgt Ion:168 Resp: 251835
Ion Ratio Lower Upper
168 100
99 66.2 42.8 64.2#



#33
1,2-Dichloroethane-d4
Concen: 68.023 ug/l
RT: 8.588 min Scan# 1111
Delta R.T. -0.000 min
Lab File: VN074805.D
Acq: 07 Oct 2022 12:42

Tgt Ion: 65 Resp: 405897
Ion Ratio Lower Upper
65 100
67 54.9 0.0 109.0





#34

1,4-Difluorobenzene

Concen: 50.000 ug/l

RT: 9.106 min Scan# 1199

Delta R.T. -0.000 min

Lab File: VN074805.D

Acq: 07 Oct 2022 12:42

Instrument :

MSVOA_N

ClientSampleId :

257248-6





Tgt Ion:114 Resp: 472108

Ion Ratio Lower Upper

114 100

63 20.9 0.0 41.2

88 16.6 0.0 30.2



#35

Dibromofluoromethane

Concen: 66.897 ug/l

RT: 8.177 min Scan# 1041

Delta R.T. 0.006 min

Lab File: VN074805.D

Acq: 07 Oct 2022 12:42




Tgt Ion:113 Resp: 319098


Ion Ratio Lower Upper

113 100

111 103.6 83.0 124.6

192 17.0 12.6 18.8









#50
Toluene-d8
Concen: 63.153 ug/l
RT: 10.571 min Scan# 1448
Delta R.T. -0.000 min
Lab File: VN074805.D
Acq: 07 Oct 2022 12:42




Instrument :
MSVOA_N
ClientSampleId :
257248-6


Tgt Ion: 98 Resp: 1292571
Ion Ratio Lower Upper
98 100
100 63.5 50.7 76.1



#62
4-Bromofluorobenzene
Concen: 52.276 ug/l
RT: 12.853 min Scan# 1836
Delta R.T. -0.000 min
Lab File: VN074805.D
Acq: 07 Oct 2022 12:42

Tgt Ion: 95 Resp: 325251
Ion Ratio Lower Upper
95 100
174 79.2 0.0 149.4
176 77.5 0.0 142.0









#63
Chlorobenzene-d5
Concen: 50.000 ug/l
RT: 11.871 min Scan# 1669
Delta R.T. -0.000 min
Lab File: VN074805.D
Acq: 07 Oct 2022 12:42

Instrument :
MSVOA_N
ClientSampleId :
257248-6

Tgt Ion:117 Resp: 462875
Ion Ratio Lower Upper
117 100
82 57.4 45.4 68.2
119 30.7 25.1 37.7



#72
1,4-Dichlorobenzene-d4
Concen: 50.000 ug/l
RT: 13.794 min Scan# 1996
Delta R.T. -0.000 min
Lab File: VN074805.D
Acq: 07 Oct 2022 12:42

Tgt Ion:152 Resp: 235753
Ion Ratio Lower Upper
152 100
115 56.8 30.0 90.0
150 153.2 0.0 356.4

