

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN100722\
 Data File : VN074812.D
 Acq On : 07 Oct 2022 15:31
 Operator : JC\MD
 Sample : N4977-01
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 251032-3

Quant Time: Oct 07 15:49:10 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N100522W.M
 Quant Title : SW846 8260
 QLast Update : Thu Oct 06 05:37:50 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.236	168	293827	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	526364	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.871	117	537512	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	257234	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.588	65	368798	52.973	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	105.940%
35) Dibromofluoromethane	8.183	113	281543	52.940	ug/l	0.01
Spiked Amount	50.000	Range	75 - 124	Recovery	=	105.880%
50) Toluene-d8	10.571	98	1147431	50.283	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	100.560%
62) 4-Bromofluorobenzene	12.853	95	298095	42.973	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	85.940%


Target Compounds	Qvalue


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN100722\
Data File : VN074812.D
Acq On : 07 Oct 2022 15:31
Operator : JC\MD
Sample : N4977-01
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 16 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
251032-3

Quant Time: Oct 07 15:49:10 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N100522W.M
Quant Title : SW846 8260
QLast Update : Thu Oct 06 05:37:50 2022
Response via : Initial Calibration









#1
Pentafluorobenzene
Concen: 50.000 ug/l
RT: 8.236 min Scan# 1051
Delta R.T. -0.000 min
Lab File: VN074812.D
Acq: 07 Oct 2022 15:31




Instrument :
MSVOA_N
ClientSampleId :
251032-3


Tgt Ion:168 Resp: 293827
Ion Ratio Lower Upper
168 100
99 58.4 42.8 64.2



#33
1,2-Dichloroethane-d4
Concen: 52.973 ug/l
RT: 8.588 min Scan# 1111
Delta R.T. -0.000 min
Lab File: VN074812.D
Acq: 07 Oct 2022 15:31

Tgt Ion: 65 Resp: 368798
Ion Ratio Lower Upper
65 100
67 54.6 0.0 109.0





#34

1,4-Difluorobenzene

Concen: 50.000 ug/l

RT: 9.106 min Scan# 1199

Delta R.T. -0.000 min

Lab File: VN074812.D

Acq: 07 Oct 2022 15:31

Instrument :

MSVOA_N

ClientSampleId :

251032-3





Tgt Ion:114 Resp: 526364

Ion Ratio Lower Upper

114 100

63 21.9 0.0 41.2

88 15.5 0.0 30.2



#35

Dibromofluoromethane

Concen: 52.940 ug/l

RT: 8.183 min Scan# 1042

Delta R.T. 0.012 min

Lab File: VN074812.D

Acq: 07 Oct 2022 15:31




Tgt Ion:113 Resp: 281543


Ion Ratio Lower Upper

113 100

111 106.4 83.0 124.6

192 16.5 12.6 18.8









#50
Toluene-d8
Concen: 50.283 ug/l
RT: 10.571 min Scan# 1448
Delta R.T. -0.000 min
Lab File: VN074812.D
Acq: 07 Oct 2022 15:31




Instrument :
MSVOA_N
ClientSampleId :
251032-3


Tgt Ion: 98 Resp: 1147431
Ion Ratio Lower Upper
98 100
100 63.9 50.7 76.1



#62
4-Bromofluorobenzene
Concen: 42.973 ug/l
RT: 12.853 min Scan# 1836
Delta R.T. -0.000 min
Lab File: VN074812.D
Acq: 07 Oct 2022 15:31

Tgt Ion: 95 Resp: 298095
Ion Ratio Lower Upper
95 100
174 80.5 0.0 149.4
176 75.0 0.0 142.0









#63
Chlorobenzene-d5
Concen: 50.000 ug/l
RT: 11.871 min Scan# 1669
Delta R.T. -0.000 min
Lab File: VN074812.D
Acq: 07 Oct 2022 15:31

Instrument :
MSVOA_N
ClientSampleId :
251032-3

Tgt Ion:117 Resp: 537512
Ion Ratio Lower Upper
117 100
82 60.6 45.4 68.2
119 31.9 25.1 37.7



#72
1,4-Dichlorobenzene-d4
Concen: 50.000 ug/l
RT: 13.794 min Scan# 1996
Delta R.T. -0.000 min
Lab File: VN074812.D
Acq: 07 Oct 2022 15:31

Tgt Ion:152 Resp: 257234
Ion Ratio Lower Upper
152 100
115 59.1 30.0 90.0
150 157.9 0.0 356.4

