

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN011023\
 Data File : VN076183.D
 Acq On : 10 Jan 2023 15:37
 Operator : JC\MD
 Sample : N6245-03MEDL 40X
 Misc : 3.60gr/5.0mL/100uL/5.0mL/MSVOA_N/MeOH
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 WASTE-VOCMEDL

Quant Time: Jan 11 01:06:21 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N121422W.M
 Quant Title : SW846 8260
 QLast Update : Wed Dec 14 13:03:41 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.230	168	329033	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	520629	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	448718	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	184949	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.583	65	174908	48.863	ug/l	0.00
Spiked Amount 50.000	Range 74 - 125		Recovery =	97.720%		
35) Dibromofluoromethane	8.171	113	155961	48.372	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery =	96.740%		
50) Toluene-d8	10.571	98	593396	49.225	ug/l	0.00
Spiked Amount 50.000	Range 86 - 113		Recovery =	98.440%		
62) 4-Bromofluorobenzene	12.847	95	190029	47.594	ug/l	0.00
Spiked Amount 50.000	Range 83 - 123		Recovery =	95.180%		
Target Compounds						Qvalue
27) cis-1,2-Dichloroethene	7.494	96	180858	47.350	ug/l	94
44) Trichloroethene	9.359	130	71287	19.149	ug/l	98
52) Toluene	10.630	92	22277	2.588	ug/l	97
67) Ethyl Benzene	11.971	91	17684	1.138	ug/l	99
68) m/p-Xylenes	12.071	106	22206	3.609	ug/l	91

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN011023\
Data File : VN076183.D
Acq On : 10 Jan 2023 15:37
Operator : JC\MD
Sample : N6245-03MEDL 40X
Misc : 3.60gr/5.0mL/100uL/5.0mL/MSVOA_N/MeOH
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
WASTE-VOCMEDL

Quant Time: Jan 11 01:06:21 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N121422W.M
Quant Title : SW846 8260
QLast Update : Wed Dec 14 13:03:41 2022
Response via : Initial Calibration

