

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN011221\
 Data File : VN065490.D
 Acq On : 12 Jan 2021 13:06
 Operator : JC/MD
 Sample : M1058-01
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB-2021-WS-000-05

Quant Time: Jan 12 23:58:41 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N122220W.M
 Quant Title : SW846 8260
 QLast Update : Tue Dec 22 15:42:18 2020
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.630	168	213280	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.553	114	349019	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.379	117	376811	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.318	152	152770	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.990	65	115523	47.98	ug/l	0.00
Spiked Amount 50.000	Range 61 - 141		Recovery = 95.96%			
35) Dibromofluoromethane	7.553	113	108003	51.29	ug/l	0.00
Spiked Amount 50.000	Range 69 - 133		Recovery = 102.58%			
50) Toluene-d8	10.061	98	439223	55.65	ug/l	0.00
Spiked Amount 50.000	Range 65 - 126		Recovery = 111.30%			
62) 4-Bromofluorobenzene	12.376	95	166123	52.90	ug/l	0.00
Spiked Amount 50.000	Range 58 - 135		Recovery = 105.80%			
Target Compounds						Qvalue
20) Methylene Chloride	4.508	84	10494	4.47	ug/l	96
43) Isopropyl Acetate	8.135	43	16278	3.00	ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN011221\
Data File : VN065490.D
Acq On : 12 Jan 2021 13:06
Operator : JC/MD
Sample : M1058-01
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB-2021-WS-000-05

Quant Time: Jan 12 23:58:41 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N122220W.M
Quant Title : SW846 8260
QLast Update : Tue Dec 22 15:42:18 2020
Response via : Initial Calibration

