

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN020223\
 Data File : VN076512.D
 Acq On : 02 Feb 2023 16:03
 Operator : JC\MD
 Sample : 01386-14
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 JPP-15-013123

Quant Time: Feb 03 01:01:53 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N011623W.M
 Quant Title : SW846 8260
 QLast Update : Tue Jan 17 03:59:22 2023
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.230	168	567627	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	938897	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	865897	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	343428	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.582	65	336981	53.021	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	= 106.040%	
35) Dibromofluoromethane	8.177	113	286600	48.798	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 97.600%	
50) Toluene-d8	10.571	98	1128210	52.014	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	= 104.020%	
62) 4-Bromofluorobenzene	12.853	95	382038	53.953	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	= 107.900%	
Target Compounds						
					Qvalue	
16) Acetone	4.447	43	35727	17.467	ug/l	95
20) Methylene Chloride	5.289	84	30802	4.562	ug/l	87
37) Ethyl Acetate	7.571	43	55271	8.287	ug/l	99
43) Isopropyl Acetate	8.694	43	256267	23.533	ug/l #	89
95) Naphthalene	15.635	128	2135	1.954	ug/l #	78

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN020223\
Data File : VN076512.D
Acq On : 02 Feb 2023 16:03
Operator : JC\MD
Sample : 01386-14
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
JPP-15-013123

Quant Time: Feb 03 01:01:53 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N011623W.M
Quant Title : SW846 8260
QLast Update : Tue Jan 17 03:59:22 2023
Response via : Initial Calibration

