

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN020223\
 Data File : VN076513.D
 Acq On : 02 Feb 2023 16:27
 Operator : JC\MD
 Sample : 01390-02
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 STOCKPILE-1

Quant Time: Feb 03 01:02:21 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N011623W.M
 Quant Title : SW846 8260
 QLast Update : Tue Jan 17 03:59:22 2023
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.235	168	596875	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	1001988	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.870	117	987902	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	424939	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.588	65	362863	54.295	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	= 108.600%	
35) Dibromofluoromethane	8.176	113	305161	48.687	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 97.380%	
50) Toluene-d8	10.570	98	1172055	50.633	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	= 101.260%	
62) 4-Bromofluorobenzene	12.853	95	465137	61.553	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	= 123.100%#	
Target Compounds						
					Qvalue	
16) Acetone	4.453	43	40431	18.804	ug/l	100
20) Methylene Chloride	5.294	84	19214	2.458	ug/l	92
37) Ethyl Acetate	7.571	43	36952	5.205	ug/l	98
43) Isopropyl Acetate	8.694	43	213404	18.363	ug/l #	86

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN020223\
Data File : VN076513.D
Acq On : 02 Feb 2023 16:27
Operator : JC\MD
Sample : 01390-02
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
STOCKPILE-1

Quant Time: Feb 03 01:02:21 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N011623W.M
Quant Title : SW846 8260
QLast Update : Tue Jan 17 03:59:22 2023
Response via : Initial Calibration

