

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN020223\
 Data File : VN076514.D
 Acq On : 02 Feb 2023 16:51
 Operator : JC\MD
 Sample : 01390-05
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 STOCKPILE-2

Quant Time: Feb 03 01:02:45 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N011623W.M
 Quant Title : SW846 8260
 QLast Update : Tue Jan 17 03:59:22 2023
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.236	168	521596	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	874864	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.871	117	802137	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	306539	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.583	65	318833	54.593	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	= 109.180%	
35) Dibromofluoromethane	8.177	113	263451	48.140	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 96.280%	
50) Toluene-d8	10.571	98	1038076	51.362	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	= 102.720%	
62) 4-Bromofluorobenzene	12.853	95	345352	52.342	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	= 104.680%	
Target Compounds						
					Qvalue	
16) Acetone	4.448	43	37921	20.187	ug/l	98
20) Methylene Chloride	5.300	84	19470	2.948	ug/l	96
37) Ethyl Acetate	7.577	43	51575	8.299	ug/l	97
43) Isopropyl Acetate	8.694	43	283404	27.930	ug/l #	86

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN020223\
Data File : VN076514.D
Acq On : 02 Feb 2023 16:51
Operator : JC\MD
Sample : 01390-05
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
STOCKPILE-2

Quant Time: Feb 03 01:02:45 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N011623W.M
Quant Title : SW846 8260
QLast Update : Tue Jan 17 03:59:22 2023
Response via : Initial Calibration

