

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN020223\
 Data File : VN076515.D
 Acq On : 02 Feb 2023 17:15
 Operator : JC\MD
 Sample : PB150614ZHE#1
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB150614ZHE#1

Quant Time: Feb 03 01:03:11 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N011623W.M
 Quant Title : SW846 8260
 QLast Update : Tue Jan 17 03:59:22 2023
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.229	168	586532	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	985448	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.870	117	901091	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	433683	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.582	65	353273	53.793	ug/l	0.00
Spiked Amount	50.000	Range 74 - 125	Recovery	= 107.580%		
35) Dibromofluoromethane	8.176	113	299614	48.604	ug/l	0.00
Spiked Amount	50.000	Range 75 - 124	Recovery	= 97.200%		
50) Toluene-d8	10.570	98	1175748	51.645	ug/l	0.00
Spiked Amount	50.000	Range 86 - 113	Recovery	= 103.300%		
62) 4-Bromofluorobenzene	12.853	95	437216	58.829	ug/l	0.00
Spiked Amount	50.000	Range 83 - 123	Recovery	= 117.660%		
Target Compounds						
16) Acetone	4.453	43	37263	17.631	ug/l	100
20) Methylene Chloride	5.288	84	23588	3.223	ug/l	96
37) Ethyl Acetate	7.571	43	52371	7.485	ug/l	100
43) Isopropyl Acetate	8.694	43	377771	33.052	ug/l #	79

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN020223\
Data File : VN076515.D
Acq On : 02 Feb 2023 17:15
Operator : JC\MD
Sample : PB150614ZHE#1
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB150614ZHE#1

Quant Time: Feb 03 01:03:11 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N011623W.M
Quant Title : SW846 8260
QLast Update : Tue Jan 17 03:59:22 2023
Response via : Initial Calibration

