

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN020223\
 Data File : VN076516.D
 Acq On : 02 Feb 2023 17:39
 Operator : JC\MD
 Sample : PB150614ZHE#2
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB150614ZHE#2

Quant Time: Feb 03 01:03:39 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N011623W.M
 Quant Title : SW846 8260
 QLast Update : Tue Jan 17 03:59:22 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.230	168	539194	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	906829	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	835367	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	330998	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.583	65	322816	53.471	ug/l	0.00
Spiked Amount	50.000	Range 74 - 125	Recovery	=	106.940%	
35) Dibromofluoromethane	8.177	113	271090	47.789	ug/l	0.00
Spiked Amount	50.000	Range 75 - 124	Recovery	=	95.580%	
50) Toluene-d8	10.571	98	1079323	51.520	ug/l	0.00
Spiked Amount	50.000	Range 86 - 113	Recovery	=	103.040%	
62) 4-Bromofluorobenzene	12.853	95	367040	53.668	ug/l	0.00
Spiked Amount	50.000	Range 83 - 123	Recovery	=	107.340%	
Target Compounds						
					Qvalue	
16) Acetone	4.453	43	39206	20.190	ug/l	99
20) Methylene Chloride	5.295	84	26398	4.056	ug/l	93
37) Ethyl Acetate	7.577	43	52114	8.091	ug/l	97
43) Isopropyl Acetate	8.694	43	417546	39.699	ug/l #	76

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN020223\
Data File : VN076516.D
Acq On : 02 Feb 2023 17:39
Operator : JC\MD
Sample : PB150614ZHE#2
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB150614ZHE#2

Quant Time: Feb 03 01:03:39 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N011623W.M
Quant Title : SW846 8260
QLast Update : Tue Jan 17 03:59:22 2023
Response via : Initial Calibration

