

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN020223\
 Data File : VN076517.D
 Acq On : 02 Feb 2023 18:03
 Operator : JC\MD
 Sample : PB150614ZHE#3
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB150614ZHE#3

Quant Time: Feb 03 01:04:04 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N011623W.M
 Quant Title : SW846 8260
 QLast Update : Tue Jan 17 03:59:22 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.230	168	513689	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	865477	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.871	117	858887	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	356152	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.588	65	304735	52.982	ug/l	0.00
Spiked Amount 50.000	Range 74 - 125		Recovery = 105.960%			
35) Dibromofluoromethane	8.171	113	257328	47.531	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery = 95.060%			
50) Toluene-d8	10.571	98	1114992	55.766	ug/l	0.00
Spiked Amount 50.000	Range 86 - 113		Recovery = 111.540%			
62) 4-Bromofluorobenzene	12.853	95	390728	59.862	ug/l	0.00
Spiked Amount 50.000	Range 83 - 123		Recovery = 119.720%			
Target Compounds					Qvalue	
16) Acetone	4.453	43	39775	21.505	ug/l	98
20) Methylene Chloride	5.283	84	26562	4.318	ug/l	89
37) Ethyl Acetate	7.571	43	51005	8.296	ug/l	97
43) Isopropyl Acetate	8.694	43	460831	45.908	ug/l #	74

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN020223\
Data File : VN076517.D
Acq On : 02 Feb 2023 18:03
Operator : JC\MD
Sample : PB150614ZHE#3
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB150614ZHE#3

Quant Time: Feb 03 01:04:04 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N011623W.M
Quant Title : SW846 8260
QLast Update : Tue Jan 17 03:59:22 2023
Response via : Initial Calibration

