

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN021924\
 Data File : VN081069.D
 Acq On : 19 Feb 2024 11:41
 Operator : JC\MD
 Sample : VN0219WBL01
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VN0219WBL01

Quant Time: Feb 20 00:19:24 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N021624W.M
 Quant Title : SW846 8260
 QLast Update : Fri Feb 16 23:45:20 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.224	168	539266	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.100	114	983304	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	855492	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	350427	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.577	65	319767	48.377	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	96.760%
35) Dibromofluoromethane	8.165	113	255110	45.881	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	91.760%
50) Toluene-d8	10.565	98	996133	50.441	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	100.880%
62) 4-Bromofluorobenzene	12.847	95	326631	44.017	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	88.040%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN021924\
Data File : VN081069.D
Acq On : 19 Feb 2024 11:41
Operator : JC\MD
Sample : VN0219WBL01
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VN0219WBL01

Quant Time: Feb 20 00:19:24 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N021624W.M
Quant Title : SW846 8260
QLast Update : Fri Feb 16 23:45:20 2024
Response via : Initial Calibration

