

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN021924\
 Data File : VN081076.D
 Acq On : 19 Feb 2024 15:18
 Operator : JC\MD
 Sample : PB159057ZHE#04
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB159057ZHE#04

Quant Time: Feb 20 00:23:00 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N021624W.M
 Quant Title : SW846 8260
 QLast Update : Fri Feb 16 23:45:20 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.224	168	400190	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.100	114	756479	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.871	117	663494	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	268584	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.583	65	284849	58.071	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	= 116.140%	
35) Dibromofluoromethane	8.171	113	226865	53.035	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 106.060%	
50) Toluene-d8	10.571	98	876435	57.686	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	= 115.380%#	
62) 4-Bromofluorobenzene	12.853	95	292081	51.163	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	= 102.320%	
Target Compounds						
					Qvalue	
16) Acetone	4.424	43	55134	28.565	ug/l	92
18) Methyl Acetate	5.024	43	5244	1.114	ug/l #	89
20) Methylene Chloride	5.277	84	30491	5.275	ug/l #	88
37) Ethyl Acetate	7.565	43	8302	1.255	ug/l #	94
43) Isopropyl Acetate	8.688	43	378253	32.699	ug/l #	87

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN021924\
Data File : VN081076.D
Acq On : 19 Feb 2024 15:18
Operator : JC\MD
Sample : PB159057ZHE#04
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB159057ZHE#04

Quant Time: Feb 20 00:23:00 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N021624W.M
Quant Title : SW846 8260
QLast Update : Fri Feb 16 23:45:20 2024
Response via : Initial Calibration

