

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN021924\
 Data File : VN081082.D
 Acq On : 19 Feb 2024 17:41
 Operator : JC\MD
 Sample : P1514-02
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 ETGI-343

Quant Time: Feb 20 00:25:29 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N021624W.M
 Quant Title : SW846 8260
 QLast Update : Fri Feb 16 23:45:20 2024
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.224	168	368207	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	707208	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.870	117	624435	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	241481	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.582	65	275695	61.087	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	= 122.180%	
35) Dibromofluoromethane	8.171	113	217742	54.448	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 108.900%	
50) Toluene-d8	10.571	98	816585	57.492	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	= 114.980%#	
62) 4-Bromofluorobenzene	12.853	95	271779	50.923	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	= 101.840%	
Target Compounds						
					Qvalue	
16) Acetone	4.436	43	71758	40.408	ug/l	99
20) Methylene Chloride	5.277	84	45426	9.128	ug/l	91
37) Ethyl Acetate	7.565	43	7124	1.152	ug/l #	88
43) Isopropyl Acetate	8.688	43	264808	24.487	ug/l #	88

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN021924\
Data File : VN081082.D
Acq On : 19 Feb 2024 17:41
Operator : JC\MD
Sample : P1514-02
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
ETGI-343

Quant Time: Feb 20 00:25:29 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N021624W.M
Quant Title : SW846 8260
QLast Update : Fri Feb 16 23:45:20 2024
Response via : Initial Calibration

