

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN022024\
 Data File : VN081094.D
 Acq On : 20 Feb 2024 13:31
 Operator : JC\MD
 Sample : P1505-12
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 WC-2S

Quant Time: Feb 20 13:48:31 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N021624W.M
 Quant Title : SW846 8260
 QLast Update : Fri Feb 16 23:45:20 2024
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.229	168	369902	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.100	114	707296	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	618982	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	250852	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.577	65	226870	50.038	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	= 100.080%	
35) Dibromofluoromethane	8.165	113	186146	46.542	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 93.080%	
50) Toluene-d8	10.571	98	642867	45.255	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	= 90.520%	
62) 4-Bromofluorobenzene	12.853	95	235311	44.085	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	= 88.160%	
Target Compounds						Qvalue
16) Acetone	4.430	43	48991	27.461	ug/l	91
20) Methylene Chloride	5.283	84	42558	8.449	ug/l #	91
43) Isopropyl Acetate	8.688	43	727554	67.270	ug/l #	71

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN022024\
Data File : VN081094.D
Acq On : 20 Feb 2024 13:31
Operator : JC\MD
Sample : P1505-12
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
WC-2S

Quant Time: Feb 20 13:48:31 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N021624W.M
Quant Title : SW846 8260
QLast Update : Fri Feb 16 23:45:20 2024
Response via : Initial Calibration

