

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN022024\
 Data File : VN081095.D
 Acq On : 20 Feb 2024 13:54
 Operator : JC\MD
 Sample : P1505-16
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 WC-2D

Quant Time: Feb 20 14:29:07 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N021624W.M
 Quant Title : SW846 8260
 QLast Update : Fri Feb 16 23:45:20 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.224	168	353655	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	671262	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	586108	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.788	152	244530	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.577	65	213041	49.146	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	98.300%
35) Dibromofluoromethane	8.165	113	179348	47.249	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	94.500%
50) Toluene-d8	10.565	98	604819	44.863	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	89.720%
62) 4-Bromofluorobenzene	12.847	95	217987	43.031	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	86.060%
Target Compounds						
					Qvalue	
16) Acetone	4.424	43	51988	30.480	ug/l	95
20) Methylene Chloride	5.283	84	44815	9.402	ug/l	91
43) Isopropyl Acetate	8.688	43	340871	33.209	ug/l #	86

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN022024\
Data File : VN081095.D
Acq On : 20 Feb 2024 13:54
Operator : JC\MD
Sample : P1505-16
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
WC-2D

Quant Time: Feb 20 14:29:07 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N021624W.M
Quant Title : SW846 8260
QLast Update : Fri Feb 16 23:45:20 2024
Response via : Initial Calibration

