

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN022024\
 Data File : VN081096.D
 Acq On : 20 Feb 2024 14:18
 Operator : JC\MD
 Sample : P1514-02
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 ETGI-343

Quant Time: Feb 21 00:19:53 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N021624W.M
 Quant Title : SW846 8260
 QLast Update : Fri Feb 16 23:45:20 2024
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.224	168	333016	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.100	114	629679	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	543507	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	228204	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.583	65	213871	52.396	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	104.800%
35) Dibromofluoromethane	8.165	113	175417	49.266	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	98.540%
50) Toluene-d8	10.565	98	599583	47.411	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	94.820%
62) 4-Bromofluorobenzene	12.847	95	218601	46.002	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	92.000%
Target Compounds						
						Qvalue
16) Acetone	4.424	43	69913	43.529	ug/l	91
20) Methylene Chloride	5.277	84	45997	10.333	ug/l	86
37) Ethyl Acetate	7.565	43	6317	1.147	ug/l #	89
43) Isopropyl Acetate	8.688	43	544009	56.499	ug/l #	72

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN022024\
Data File : VN081096.D
Acq On : 20 Feb 2024 14:18
Operator : JC\MD
Sample : P1514-02
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
ETGI-343

Quant Time: Feb 21 00:19:53 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N021624W.M
Quant Title : SW846 8260
QLast Update : Fri Feb 16 23:45:20 2024
Response via : Initial Calibration

