

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN022024\
 Data File : VN081098.D
 Acq On : 20 Feb 2024 15:05
 Operator : JC\MD
 Sample : P1524-06
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 20240766-AMENDED-COMP-39N-N-3-01

Quant Time: Feb 21 00:50:56 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N021624W.M
 Quant Title : SW846 8260
 QLast Update : Fri Feb 16 23:45:20 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.224	168	339436	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	642719	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	540740	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	226288	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.577	65	187451	45.055	ug/l	0.00
Spiked Amount 50.000	Range 74 - 125		Recovery =	90.100%		
35) Dibromofluoromethane	8.224	113	932	0.256	ug/l	0.06
Spiked Amount 50.000	Range 75 - 124		Recovery =	0.520%#		
50) Toluene-d8	10.571	98	533186	41.306	ug/l	0.00
Spiked Amount 50.000	Range 86 - 113		Recovery =	82.620%#		
62) 4-Bromofluorobenzene	12.847	95	196896	40.594	ug/l	0.00
Spiked Amount 50.000	Range 83 - 123		Recovery =	81.180%#		
Target Compounds						Qvalue
16) Acetone	4.430	43	48990	29.925	ug/l	96
20) Methylene Chloride	5.277	84	15493	2.780	ug/l #	79

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN022024\
Data File : VN081098.D
Acq On : 20 Feb 2024 15:05
Operator : JC\MD
Sample : P1524-06
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
20240766-AMENDED-COMP-39N-N-3-01

Quant Time: Feb 21 00:50:56 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N021624W.M
Quant Title : SW846 8260
QLast Update : Fri Feb 16 23:45:20 2024
Response via : Initial Calibration

