

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN022024\
 Data File : VN081099.D
 Acq On : 20 Feb 2024 15:29
 Operator : JC\MD
 Sample : P1524-07
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 20240767-AMENDED-COMP-39N-N-3-02

Quant Time: Feb 21 00:20:59 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N021624W.M
 Quant Title : SW846 8260
 QLast Update : Fri Feb 16 23:45:20 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.224	168	353565	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.100	114	667245	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	564462	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	241770	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.582	65	212893	49.125	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	98.240%
35) Dibromofluoromethane	8.165	113	61	0.016	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	0.040%#
50) Toluene-d8	10.571	98	597392	44.578	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	89.160%
62) 4-Bromofluorobenzene	12.847	95	221298	43.948	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	87.900%
Target Compounds						
					Qvalue	
16) Acetone	4.430	43	53966	31.647	ug/l	93
20) Methylene Chloride	5.283	84	15614	2.659	ug/l #	81

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN022024\
Data File : VN081099.D
Acq On : 20 Feb 2024 15:29
Operator : JC\MD
Sample : P1524-07
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 15 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
20240767-AMENDED-COMP-39N-N-3-02

Quant Time: Feb 21 00:20:59 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N021624W.M
Quant Title : SW846 8260
QLast Update : Fri Feb 16 23:45:20 2024
Response via : Initial Calibration

