

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN022024\
 Data File : VN081100.D
 Acq On : 20 Feb 2024 15:53
 Operator : JC\MD
 Sample : P1524-08
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 20240768-AMENDED-COMP-39N-N-3-03

Quant Time: Feb 21 00:21:18 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N021624W.M
 Quant Title : SW846 8260
 QLast Update : Fri Feb 16 23:45:20 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.224	168	336122	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	629583	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.870	117	533820	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	222524	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.582	65	185571	45.042	ug/l	0.00
Spiked Amount 50.000	Range 74 - 125		Recovery =	90.080%		
35) Dibromofluoromethane	8.229	113	516	0.145	ug/l	0.06
Spiked Amount 50.000	Range 75 - 124		Recovery =	0.280%#		
50) Toluene-d8	10.570	98	525220	41.537	ug/l	0.00
Spiked Amount 50.000	Range 86 - 113		Recovery =	83.080%#		
62) 4-Bromofluorobenzene	12.853	95	195332	41.112	ug/l	0.00
Spiked Amount 50.000	Range 83 - 123		Recovery =	82.220%#		
Target Compounds						Qvalue
16) Acetone	4.424	43	47516	29.311	ug/l	94
20) Methylene Chloride	5.277	84	16358	3.027	ug/l #	75

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN022024\
Data File : VN081100.D
Acq On : 20 Feb 2024 15:53
Operator : JC\MD
Sample : P1524-08
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 16 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
20240768-AMENDED-COMP-39N-N-3-03

Quant Time: Feb 21 00:21:18 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N021624W.M
Quant Title : SW846 8260
QLast Update : Fri Feb 16 23:45:20 2024
Response via : Initial Calibration

