

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN022024\
 Data File : VN081103.D
 Acq On : 20 Feb 2024 17:04
 Operator : JC\MD
 Sample : P1524-08RE
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 20240768-AMENDED-COMP-39N-N-3-03RE

Quant Time: Feb 21 00:22:24 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N021624W.M
 Quant Title : SW846 8260
 QLast Update : Fri Feb 16 23:45:20 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.224	168	361201	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.100	114	675069	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	568940	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	247185	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.576	65	212855	48.078	ug/l	0.00
Spiked Amount 50.000	Range 74 - 125		Recovery =	96.160%		
35) Dibromofluoromethane	8.171	113	2445	0.641	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery =	1.280%#		
50) Toluene-d8	10.570	98	595249	43.904	ug/l	0.00
Spiked Amount 50.000	Range 86 - 113		Recovery =	87.800%		
62) 4-Bromofluorobenzene	12.853	95	223042	43.781	ug/l	0.00
Spiked Amount 50.000	Range 83 - 123		Recovery =	87.560%		
Target Compounds						Qvalue
16) Acetone	4.430	43	48208	27.673	ug/l	95
20) Methylene Chloride	5.277	84	15718	2.607	ug/l #	88

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN022024\
Data File : VN081103.D
Acq On : 20 Feb 2024 17:04
Operator : JC\MD
Sample : P1524-08RE
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 19 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
20240768-AMENDED-COMP-39N-N-3-03RE

Quant Time: Feb 21 00:22:24 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N021624W.M
Quant Title : SW846 8260
QLast Update : Fri Feb 16 23:45:20 2024
Response via : Initial Calibration

