

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN022024\
 Data File : VN081106.D
 Acq On : 20 Feb 2024 18:16
 Operator : JC\MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 22 Sample Multiplier: 1

Instrument :
 MSVOA_N
 LabSampleId :
 VSTDCCC050

Quant Time: Feb 21 00:24:21 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N021624W.M
 Quant Title : SW846 8260
 QLast Update : Fri Feb 16 23:45:20 2024
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	1.000	1.000	0.0	91	0.00
2 T	Dichlorodifluoromethane	0.647	0.692	-7.0	92	0.00
3 P	Chloromethane	0.833	0.648	22.2	83	0.00
4 C	Vinyl Chloride	0.718	0.709	1.3#	92	0.00
5 T	Bromomethane	0.510	0.477	6.5	89	0.00
6 T	Chloroethane	0.556	0.475	14.6	93	0.00
7 T	Trichlorofluoromethane	1.049	1.069	-1.9	94	0.00
8 T	Diethyl Ether	0.384	0.396	-3.1	94	0.00
9 T	1,1,2-Trichlorotrifluoroeth	0.585	0.592	-1.2	93	0.00
10 T	Methyl Iodide	0.653	0.735	-12.6	99	0.00
11 T	Tert butyl alcohol	0.113	0.132	-16.8	118	0.00
12 CM	1,1-Dichloroethene	0.565	0.582	-3.0#	96	0.00
13 T	Acrolein	0.060	0.080	-33.3#	133	0.00
14 T	Allyl chloride	0.839	0.847	-1.0	94	0.01
15 T	Acrylonitrile	0.294	0.309	-5.1	100	0.00
16 T	Acetone	0.241	0.189	21.6	77	0.00
17 T	Carbon Disulfide	1.577	1.488	5.6	90	0.00
18 T	Methyl Acetate	0.588	0.703	-19.6	108	0.00
19 T	Methyl tert-butyl Ether	1.937	2.027	-4.6	96	0.00
20 T	Methylene Chloride	0.693	0.652	5.9	96	0.00
21 T	trans-1,2-Dichloroethene	0.631	0.619	1.9	94	0.00
22 T	Diisopropyl ether	1.903	2.012	-5.7	97	0.00
23 T	Vinyl Acetate	1.544	1.574	-1.9	92	0.00
24 P	1,1-Dichloroethane	1.100	1.156	-5.1	97	0.00
25 T	2-Butanone	0.390	0.373	4.4	92	0.00
26 T	2,2-Dichloropropane	1.031	1.028	0.3	90	0.00
27 T	cis-1,2-Dichloroethene	0.733	0.742	-1.2	94	0.00
28 T	Bromochloromethane	0.456	0.460	-0.9	94	0.00
29 T	Tetrahydrofuran	0.260	0.265	-1.9	98	0.00
30 C	Chloroform	1.170	1.214	-3.8#	97	0.00
31 T	Cyclohexane	1.058	1.004	5.1	95	0.00
32 T	1,1,1-Trichloroethane	1.065	1.102	-3.5	96	0.00
33 S	1,2-Dichloroethane-d4	0.613	0.550	10.3	87	0.00
34 I	1,4-Difluorobenzene	1.000	1.000	0.0	90	0.00
35 S	Dibromofluoromethane	0.283	0.263	7.1	86	0.00
36 T	1,1-Dichloropropene	0.484	0.489	-1.0	94	0.00
37 T	Ethyl Acetate	0.437	0.444	-1.6	94	0.00
38 T	Carbon Tetrachloride	0.490	0.527	-7.6	95	0.00
39 T	Methylcyclohexane	0.565	0.588	-4.1	94	0.00
40 TM	Benzene	1.402	1.472	-5.0	96	0.00
41 T	Methacrylonitrile	0.249	0.260	-4.4	105	0.00
42 TM	1,2-Dichloroethane	0.489	0.508	-3.9	96	0.00
43 T	Isopropyl Acetate	0.765	0.815	-6.5	98	0.00
44 TM	Trichloroethene	0.373	0.384	-2.9	98	0.00
45 C	1,2-Dichloropropane	0.351	0.369	-5.1#	97	0.00
46 T	Dibromomethane	0.256	0.261	-2.0	96	0.00
47 T	Bromodichloromethane	0.484	0.526	-8.7	95	0.00
48 T	Methyl methacrylate	0.356	0.366	-2.8	97	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN022024\
 Data File : VN081106.D
 Acq On : 20 Feb 2024 18:16
 Operator : JC\MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 22 Sample Multiplier: 1

Instrument :
 MSVOA_N
 LabSampleId :
 VSTDCCC050

Quant Time: Feb 21 00:24:21 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N021624W.M
 Quant Title : SW846 8260
 QLast Update : Fri Feb 16 23:45:20 2024
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
49 T	1,4-Dioxane	0.008	0.007	12.5	96	0.00
50 S	Toluene-d8	1.004	0.869	13.4	83	0.00
51 T	4-Methyl-2-Pentanone	0.440	0.472	-7.3	98	0.00
52 CM	Toluene	0.888	0.941	-6.0#	95	0.00
53 T	t-1,3-Dichloropropene	0.538	0.592	-10.0	94	0.00
54 T	cis-1,3-Dichloropropene	0.581	0.618	-6.4	95	0.00
55 T	1,1,2-Trichloroethane	0.341	0.366	-7.3	95	0.00
56 T	Ethyl methacrylate	0.522	0.574	-10.0	97	0.00
57 T	1,3-Dichloropropane	0.590	0.632	-7.1	98	0.00
58 T	2-Chloroethyl Vinyl ether	0.262	0.249	5.0	87	0.00
59 T	2-Hexanone	0.349	0.347	0.6	86	0.00
60 T	Dibromochloromethane	0.363	0.401	-10.5	94	0.00
61 T	1,2-Dibromoethane	0.359	0.377	-5.0	95	0.00
62 S	4-Bromofluorobenzene	0.377	0.347	8.0	87	0.00
63 I	Chlorobenzene-d5	1.000	1.000	0.0	92	0.00
64 T	Tetrachloroethene	0.341	0.351	-2.9	102	0.00
65 PM	Chlorobenzene	1.082	1.101	-1.8	96	0.00
66 T	1,1,1,2-Tetrachloroethane	0.389	0.410	-5.4	96	0.00
67 C	Ethyl Benzene	1.936	1.989	-2.7#	94	0.00
68 T	m/p-Xylenes	0.742	0.762	-2.7	94	0.00
69 T	o-Xylene	0.736	0.750	-1.9	94	0.00
70 T	Styrene	1.166	1.223	-4.9	94	0.00
71 P	Bromoform	0.247	0.285	-15.4	98	0.00
72 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	99	0.00
73 T	Isopropylbenzene	4.375	4.262	2.6	96	0.00
74 T	N-amyl acetate	1.849	1.783	3.6	94	0.00
75 P	1,1,2,2-Tetrachloroethane	1.230	1.211	1.5	100	0.00
76 T	1,2,3-Trichloropropane	1.136	1.101	3.1	102	0.00
77 T	Bromobenzene	0.975	0.949	2.7	99	0.00
78 T	n-propylbenzene	5.036	5.020	0.3	96	0.00
79 T	2-Chlorotoluene	3.089	2.974	3.7	99	0.00
80 T	1,3,5-Trimethylbenzene	3.547	3.573	-0.7	97	0.00
81 T	trans-1,4-Dichloro-2-butene	0.441	0.446	-1.1	98	0.00
82 T	4-Chlorotoluene	3.020	2.953	2.2	98	0.00
83 T	tert-Butylbenzene	3.067	3.044	0.7	98	0.00
84 T	1,2,4-Trimethylbenzene	3.534	3.560	-0.7	98	0.00
85 T	sec-Butylbenzene	4.264	4.385	-2.8	98	0.00
86 T	p-Isopropyltoluene	3.542	3.673	-3.7	100	0.00
87 T	1,3-Dichlorobenzene	1.843	1.827	0.9	102	0.00
88 T	1,4-Dichlorobenzene	1.811	1.792	1.0	100	0.00
89 T	n-Butylbenzene	3.085	3.224	-4.5	100	0.00
90 T	Hexachloroethane	0.585	0.634	-8.4	100	0.00
91 T	1,2-Dichlorobenzene	1.778	1.730	2.7	99	0.00
92 T	1,2-Dibromo-3-Chloropropane	0.256	0.246	3.9	97	0.00
93 T	1,2,4-Trichlorobenzene	0.992	1.029	-3.7	101	0.00
94 T	Hexachlorobutadiene	0.370	0.374	-1.1	100	0.00
95 T	Naphthalene	3.634	3.660	-0.7	99	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN022024\
Data File : VN081106.D
Acq On : 20 Feb 2024 18:16
Operator : JC\MD
Sample : VSTDCCC050
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 22 Sample Multiplier: 1

Instrument :
MSVOA_N
LabSampleId :
VSTDCCC050

Quant Time: Feb 21 00:24:21 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N021624W.M
Quant Title : SW846 8260
QLast Update : Fri Feb 16 23:45:20 2024
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
96 T	1,2,3-Trichlorobenzene	0.984	1.018	-3.5	101	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 6