

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN032224\
 Data File : VN081536.D
 Acq On : 22 Mar 2024 10:05
 Operator : JC\MD
 Sample : VN0322MBL01
 Misc : 5.00g/10mL/100uL/5.00mL/MSVOA_N/MEOH
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VN0322MBL01

Quant Time: Mar 23 00:09:44 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N031824W.M
 Quant Title : SW846 8260
 QLast Update : Tue Mar 19 06:00:00 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.224	168	417805	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	9.100	114	713764	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	624180	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.788	152	261990	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.577	65	256799	48.197	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	96.400%
35) Dibromofluoromethane	8.165	113	220699	48.994	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	97.980%
50) Toluene-d8	10.565	98	838396	51.487	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	102.980%
62) 4-Bromofluorobenzene	12.847	95	262529	45.755	ug/l	0.00
Spiked Amount	50.000	Range	64 - 133	Recovery	=	91.520%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN032224\
Data File : VN081536.D
Acq On : 22 Mar 2024 10:05
Operator : JC\MD
Sample : VN0322MBL01
Misc : 5.00g/10mL/100uL/5.00mL/MSVOA_N/MEOH
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VN0322MBL01

Quant Time: Mar 23 00:09:44 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N031824W.M
Quant Title : SW846 8260
QLast Update : Tue Mar 19 06:00:00 2024
Response via : Initial Calibration

