

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN032224\
 Data File : VN081542.D
 Acq On : 22 Mar 2024 12:52
 Operator : JC\MD
 Sample : P1798-06DL 50X
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 S-6DL

Quant Time: Mar 23 00:12:49 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N031824W.M
 Quant Title : SW846 8260
 QLast Update : Tue Mar 19 06:00:00 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.224	168	477866	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	9.100	114	837484	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	748755	50.000	ug/l	# 0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	316648	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.577	65	303431	49.791	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	99.580%
35) Dibromofluoromethane	8.165	113	253716	48.003	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	96.000%
50) Toluene-d8	10.565	98	988632	51.745	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	103.480%
62) 4-Bromofluorobenzene	12.847	95	322139	47.850	ug/l	0.00
Spiked Amount	50.000	Range	64 - 133	Recovery	=	95.700%
Target Compounds						
						Qvalue
40) Benzene	8.606	78	922224	35.464	ug/l	94
52) Toluene	10.629	92	287111	17.371	ug/l	99
68) m/p-Xylenes	12.071	106	43727	3.794	ug/l	92
69) o-Xylene	12.400	106	24826	2.194	ug/l	97
95) Naphthalene	15.641	128	253964	11.496	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN032224\
Data File : VN081542.D
Acq On : 22 Mar 2024 12:52
Operator : JC\MD
Sample : P1798-06DL 50X
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
S-6DL

Quant Time: Mar 23 00:12:49 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N031824W.M
Quant Title : SW846 8260
QLast Update : Tue Mar 19 06:00:00 2024
Response via : Initial Calibration

