

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN032323\
 Data File : VN077141.D
 Acq On : 23 Mar 2023 11:41
 Operator : JC\MD
 Sample : VN0323MBL01
 Misc : 5.00g/10mL/100uL/5.0mL/MSVOA_N/MEOH
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VN0323MBL01

Quant Time: Mar 24 01:33:04 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N031723W.M
 Quant Title : SW846 8260
 QLast Update : Mon Mar 20 08:50:30 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.236	168	374652	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.107	114	652380	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.871	117	571064	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.795	152	212970	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.583	65	289331	48.348	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	96.700%
35) Dibromofluoromethane	8.172	113	217687	47.647	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	95.300%
50) Toluene-d8	10.571	98	805106	49.332	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	98.660%
62) 4-Bromofluorobenzene	12.854	95	239647	41.556	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	83.120%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN032323\
Data File : VN077141.D
Acq On : 23 Mar 2023 11:41
Operator : JC\MD
Sample : VN0323MBL01
Misc : 5.00g/10mL/100uL/5.0mL/MSVOA_N/MEOH
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VN0323MBL01

Quant Time: Mar 24 01:33:04 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N031723W.M
Quant Title : SW846 8260
QLast Update : Mon Mar 20 08:50:30 2023
Response via : Initial Calibration

