

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN032323\
 Data File : VN077145.D
 Acq On : 23 Mar 2023 13:29
 Operator : JC\MD
 Sample : 01849-04DL 2X
 Misc : 10.42g/10mL/100uL/5.0mL/MSVOA_N/MEOH
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 CB-10-22.0-030823DL

Quant Time: Mar 24 01:34:04 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N031723W.M
 Quant Title : SW846 8260
 QLast Update : Mon Mar 20 08:50:30 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.230	168	346322	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.101	114	625955	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.871	117	560274	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.795	152	239911	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.583	65	262758	47.499	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	95.000%
35) Dibromofluoromethane	8.172	113	202062	46.095	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	92.180%
50) Toluene-d8	10.572	98	764175	48.801	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	97.600%
62) 4-Bromofluorobenzene	12.854	95	248537	44.917	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	89.840%
Target Compounds						
44) Trichloroethene	9.354	130	169698	35.579	ug/l	Qvalue 99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN032323\
Data File : VN077145.D
Acq On : 23 Mar 2023 13:29
Operator : JC\MD
Sample : 01849-04DL 2X
Misc : 10.42g/10mL/100uL/5.0mL/MSVOA_N/MEOH
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
CB-10-22.0-030823DL

Quant Time: Mar 24 01:34:04 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N031723W.M
Quant Title : SW846 8260
QLast Update : Mon Mar 20 08:50:30 2023
Response via : Initial Calibration

