

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN032323\
 Data File : VN077146.D
 Acq On : 23 Mar 2023 13:53
 Operator : JC\MD
 Sample : 01849-05RE
 Misc : 13.21g/10mL/100uL/5.0mL/MSVOA_N/MEOH
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 CB-10-25.0-030823RE

Quant Time: Mar 24 01:34:14 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N031723W.M
 Quant Title : SW846 8260
 QLast Update : Mon Mar 20 08:50:30 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.225	168	371533	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.101	114	638618	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.872	117	574764	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.795	152	232164	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.578	65	275622	46.443	ug/l	0.00
Spiked Amount 50.000	Range 74 - 125		Recovery =	92.880%		
35) Dibromofluoromethane	8.166	113	204393	45.702	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery =	91.400%		
50) Toluene-d8	10.572	98	793036	49.640	ug/l	0.00
Spiked Amount 50.000	Range 86 - 113		Recovery =	99.280%		
62) 4-Bromofluorobenzene	12.854	95	238211	42.197	ug/l	0.00
Spiked Amount 50.000	Range 83 - 123		Recovery =	84.400%		
Target Compounds						Qvalue
18) Methyl Acetate	5.048	43	15010	1.505	ug/l	99
20) Methylene Chloride	5.260	84	15122	2.635	ug/l	92
44) Trichloroethene	9.354	130	170872	35.115	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN032323\
Data File : VN077146.D
Acq On : 23 Mar 2023 13:53
Operator : JC\MD
Sample : 01849-05RE
Misc : 13.21g/10mL/100uL/5.0mL/MSVOA_N/MEOH
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
CB-10-25.0-030823RE

Quant Time: Mar 24 01:34:14 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N031723W.M
Quant Title : SW846 8260
QLast Update : Mon Mar 20 08:50:30 2023
Response via : Initial Calibration

