

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN032323\
 Data File : VN077155.D
 Acq On : 23 Mar 2023 17:30
 Operator : JC\MD
 Sample : 02048-08
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 TRIP-BLANK

Quant Time: Mar 24 01:36:07 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N031723W.M
 Quant Title : SW846 8260
 QLast Update : Mon Mar 20 08:50:30 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.231	168	364228	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.107	114	653864	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.871	117	571955	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.795	152	211374	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.589	65	288379	49.568	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	99.140%
35) Dibromofluoromethane	8.178	113	215878	47.144	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	94.280%
50) Toluene-d8	10.572	98	805622	49.252	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	98.500%
62) 4-Bromofluorobenzene	12.854	95	232991	40.310	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	80.620%#
Target Compounds						
					Qvalue	
20) Methylene Chloride	5.295	84	3666	0.652	ug/l #	64

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN032323\
Data File : VN077155.D
Acq On : 23 Mar 2023 17:30
Operator : JC\MD
Sample : 02048-08
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 17 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
TRIP-BLANK

Quant Time: Mar 24 01:36:07 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N031723W.M
Quant Title : SW846 8260
QLast Update : Mon Mar 20 08:50:30 2023
Response via : Initial Calibration

