

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN032323\
 Data File : VN077156.D
 Acq On : 23 Mar 2023 17:55
 Operator : JC\MD
 Sample : 02048-01
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 MW-1R

Quant Time: Mar 24 01:36:16 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N031723W.M
 Quant Title : SW846 8260
 QLast Update : Mon Mar 20 08:50:30 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.236	168	364363	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.107	114	642964	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.872	117	568657	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.795	152	216113	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.589	65	288272	49.531	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	99.060%
35) Dibromofluoromethane	8.178	113	212832	47.267	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	94.540%
50) Toluene-d8	10.572	98	797762	49.598	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	99.200%
62) 4-Bromofluorobenzene	12.854	95	235056	41.356	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	82.720%#
Target Compounds						
19) Methyl tert-butyl Ether	5.807	73	58528	3.686	ug/l	Qvalue 94

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN032323\
Data File : VN077156.D
Acq On : 23 Mar 2023 17:55
Operator : JC\MD
Sample : 02048-01
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 18 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
MW-1R

Quant Time: Mar 24 01:36:16 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N031723W.M
Quant Title : SW846 8260
QLast Update : Mon Mar 20 08:50:30 2023
Response via : Initial Calibration

