

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN032323\
 Data File : VN077160.D
 Acq On : 23 Mar 2023 19:32
 Operator : JC\MD
 Sample : 02048-05
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 22 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 MW-4

Quant Time: Mar 24 01:37:08 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N031723W.M
 Quant Title : SW846 8260
 QLast Update : Mon Mar 20 08:50:30 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.231	168	354484	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.107	114	614805	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.871	117	542301	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.795	152	198620	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.583	65	277009	48.922	ug/l	0.00
Spiked Amount 50.000	Range 74 - 125		Recovery =	97.840%		
35) Dibromofluoromethane	8.172	113	207615	48.220	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery =	96.440%		
50) Toluene-d8	10.572	98	753992	49.024	ug/l	0.00
Spiked Amount 50.000	Range 86 - 113		Recovery =	98.040%		
62) 4-Bromofluorobenzene	12.854	95	218472	40.199	ug/l	0.00
Spiked Amount 50.000	Range 83 - 123		Recovery =	80.400%#		
Target Compounds						
19) Methyl tert-butyl Ether	5.807	73	1208536	78.243	ug/l	Qvalue 96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN032323\
Data File : VN077160.D
Acq On : 23 Mar 2023 19:32
Operator : JC\MD
Sample : 02048-05
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 22 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
MW-4

Quant Time: Mar 24 01:37:08 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N031723W.M
Quant Title : SW846 8260
QLast Update : Mon Mar 20 08:50:30 2023
Response via : Initial Calibration

