

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN041822\
 Data File : VN072034.D
 Acq On : 18 Apr 2022 14:42
 Operator : JC\MD
 Sample : N2441-02
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 TT-041-IDWGW-20220413

Quant Time: Apr 18 15:26:29 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\624N032422W.M
 Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
 QLast Update : Thu Mar 24 12:03:44 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Bromochloromethane	7.657	128	84201	30.000	ug/l	0.00
28) 1,4-Difluorobenzene	8.963	114	437146	30.000	ug/l	0.00
57) Chlorobenzene-d5	11.739	117	368468	30.000	ug/l	0.00
System Monitoring Compounds						
27) 1,2-Dichloroethane-d4	8.433	65	166321	28.036	ug/l	0.00
Spiked Amount	30.000	Range	91 - 110	Recovery	=	93.467%
60) 4-Bromofluorobenzene	12.727	95	125845	23.034	ug/l	0.00
Spiked Amount	30.000	Range	63 - 112	Recovery	=	76.767%
63) Toluene-d8	10.439	98	507140	30.320	ug/l	0.00
Spiked Amount	30.000	Range	91 - 112	Recovery	=	101.067%
Target Compounds						
15) Acetone	4.293	58	6598	7.363	ug/l	Qvalue # 47
37) Trichloroethene	9.210	130	43077	7.594	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN041822\
Data File : VN072034.D
Acq On : 18 Apr 2022 14:42
Operator : JC\MD
Sample : N2441-02
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
TT-041-IDWGW-20220413

Quant Time: Apr 18 15:26:29 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\624N032422W.M
Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
QLast Update : Thu Mar 24 12:03:44 2022
Response via : Initial Calibration

